

Lb13

Test results Aquakem 7.2AQ1 Page:

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/22/2025 12:21 Reviewed by : RH Instrument ID : Konelab

Test: Ammonia-N

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.050	0.0	0.223	
ICB1	0.019	0.0	0.024	
CCV1	1.035	0.0	0.221	
CCB1	0.016	0.0	0.023	
RL CHECK	0.115	0.0	0.042	
PB168128BL	0.016	0.0	0.023	
PB168128BS	1.018	0.0	0.217	
Q2085-01	0.251	0.0	0.069	
Q2085-01DUP	0.251	0.0	0.069	
Q2085-01MS	1.197	0.0	0.252	
Q2085-01MSD	1.245	0.0	0.261	
Q2085-02	0.181	0.0	0.055	
Q2085-03	0.048	0.0	0.029	
Q2085-07	0.125	0.0	0.044	
CCV2	1.000	0.0	0.214	
CCB2	0.016	0.0	0.023	
Q2085-08	0.244	0.0	0.067	
PB168122BL	0.018	0.0	0.024	
PB168122BS	1.025	0.0	0.218	
Q2088-01	0.123	0.0	0.044	
Q2093-14	0.292	0.0	0.077	
Q2093-14DUP	0.289	0.0	0.076	
Q2093-14MS	1.269	0.0	0.266	
Q2093-14MSD	1.305	0.0	0.273	
Q2102-03	0.029	0.0	0.026	
CCV3	1.070	0.0	0.227	
CCB3	0.018	0.0	0.024	

115% (50-150)
05/22/2025
RH

N 27
Mean 0.491
SD 0.5045
CV% 102.67

Aquakem v. 7.2AQ1

Results from time period:

Thu May 22 10:54:10 2025

Thu May 22 12:16:34 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Ammonia-† P		0.0189	mg/l	5/22/2025 10:54:10	
0.1PPM	A	Ammonia-† P		0.1057	mg/l	5/22/2025 10:54:11	
0.2PPM	A	Ammonia-† P		0.2045	mg/l	5/22/2025 10:54:12	
0.4PPM	A	Ammonia-† P		0.3934	mg/l	5/22/2025 10:54:13	
1.0PPM	A	Ammonia-† P		0.9839	mg/l	5/22/2025 10:54:14	
1.3PPM	A	Ammonia-† P		1.2882	mg/l	5/22/2025 10:54:15	
2.0PPM	A	Ammonia-† P		2.0387	mg/l	5/22/2025 10:54:16	
ICV1	S	Ammonia-† P		1.0496	mg/l	5/22/2025 11:50:15	
ICB1	S	Ammonia-† P		0.0187	mg/l	5/22/2025 11:50:16	
CCV1	S	Ammonia-† P		1.0352	mg/l	5/22/2025 11:50:19	
CCB1	S	Ammonia-† P		0.0165	mg/l	5/22/2025 11:50:21	
RL CHECK	S	Ammonia-† P		0.1149	mg/l	5/22/2025 11:50:23	
PB168128BL	S	Ammonia-† P		0.0164	mg/l	5/22/2025 11:50:24	
PB168128BS	S	Ammonia-† P		1.0185	mg/l	5/22/2025 12:00:59	
Q2085-01	S	Ammonia-† P		0.2514	mg/l	5/22/2025 12:01:00	
Q2085-01DUP	S	Ammonia-† P		0.2508	mg/l	5/22/2025 12:01:01	
Q2085-01MS	S	Ammonia-† P		1.1969	mg/l	5/22/2025 12:01:02	
Q2085-01MSD	S	Ammonia-† P		1.2449	mg/l	5/22/2025 12:01:03	
Q2085-02	S	Ammonia-† P		0.1814	mg/l	5/22/2025 12:01:04	
Q2085-03	S	Ammonia-† P		0.0484	mg/l	5/22/2025 12:01:05	
Q2085-07	S	Ammonia-† P		0.1253	mg/l	5/22/2025 12:01:06	
CCV2	S	Ammonia-† P		0.9998	mg/l	5/22/2025 12:01:07	
CCB2	S	Ammonia-† P		0.0164	mg/l	5/22/2025 12:01:08	
Q2085-08	S	Ammonia-† P		0.244	mg/l	5/22/2025 12:01:09	
PB168122BL	S	Ammonia-† P		0.0179	mg/l	5/22/2025 12:11:41	
PB168122BS	S	Ammonia-† P		1.0247	mg/l	5/22/2025 12:11:42	
Q2088-01	S	Ammonia-† P		0.1228	mg/l	5/22/2025 12:11:43	
Q2093-14	S	Ammonia-† P		0.2924	mg/l	5/22/2025 12:11:44	
Q2093-14DUP	S	Ammonia-† P		0.289	mg/l	5/22/2025 12:11:45	
Q2093-14MS	S	Ammonia-† P		1.2694	mg/l	5/22/2025 12:11:46	
Q2093-14MSD	S	Ammonia-† P		1.3046	mg/l	5/22/2025 12:11:47	
Q2102-03	S	Ammonia-† P		0.0289	mg/l	5/22/2025 12:11:48	
CCV3	S	Ammonia-† P		1.0697	mg/l	5/22/2025 12:11:49	
CCB3	S	Ammonia-† P		0.0185	mg/l	5/22/2025 12:11:50	

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/22/2025 10:55

Reviewed by : RM

Instrument ID : Konelab


Test Ammonia-N

Accepted 5/22/2025 10:55

Factor 5.165
Bias 0.020

Coeff. of det. 0.998740

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors ^{Re}
1	0.00PPM	0.024	0.0189	0.0000	-
2	NH3-2PPM	0.041	0.1057	0.1000	5.7
3	NH3-2PPM	0.060	0.2045	0.2000	2.2
4	NH3-2PPM	0.096	0.3934	0.4000	-1.7
5	NH3-2PPM	0.211	0.9839	1.0000	-1.6
6	NH3-2PPM	0.270	1.2882	1.3333	-0.9
7	NH3-2PPM	0.415	2.0387	2.0000	1.9

05/22/2025
RM