

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Environmental, Inc **SDG No.:** Q2090  
**Contract:** TULL01 **Lab Code:** CHEM **Case No.:** Q2090 **SAS No.:** Q2090  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Arsenic  | 1020           | 1000       | 102           | 95 - 105                  | P | 05/21/2025       | 13:19            | LB135868      |
|           | Cadmium  | 510            | 510        | 100           | 95 - 105                  | P | 05/21/2025       | 13:19            | LB135868      |
|           | Chromium | 514            | 520        | 99            | 95 - 105                  | P | 05/21/2025       | 13:19            | LB135868      |
|           | Copper   | 514            | 510        | 101           | 95 - 105                  | P | 05/21/2025       | 13:19            | LB135868      |
|           | Iron     | 10400          | 10000      | 104           | 95 - 105                  | P | 05/21/2025       | 13:19            | LB135868      |
|           | Lead     | 989            | 1000       | 99            | 95 - 105                  | P | 05/21/2025       | 13:19            | LB135868      |
|           | Selenium | 1050           | 1000       | 105           | 95 - 105                  | P | 05/21/2025       | 13:19            | LB135868      |

---



- 2a -

|                                       |                                 |                  |              |                                                            |
|---------------------------------------|---------------------------------|------------------|--------------|------------------------------------------------------------|
| <b>Client:</b>                        | <u>Tully Environmental, Inc</u> | <b>SDG No.:</b>  | <u>Q2090</u> |                                                            |
| <b>Contract:</b>                      | <u>TULL01</u>                   | <b>Lab Code:</b> | <u>CHEM</u>  | <b>Case No.:</b> <u>Q2090</u> <b>SAS No.:</b> <u>Q2090</u> |
| <b>Initial Calibration Source:</b>    | <u>EPA</u>                      |                  |              |                                                            |
| <b>Continuing Calibration Source:</b> | Inorganic Ventures              |                  |              |                                                            |

| Sample ID | Analyte  | Result | True Value | % Recovery | Acceptance Window (%R) | M | Analysis Date | Analysis Time | Run Number |
|-----------|----------|--------|------------|------------|------------------------|---|---------------|---------------|------------|
|           |          | ug/L   |            |            |                        |   |               |               |            |
| LLICV01   | Arsenic  | 17.0   | 20.0       | 85         | 80 - 120               | P | 05/21/2025    | 13:29         | LB135868   |
|           | Cadmium  | 6.06   | 6.0        | 101        | 80 - 120               | P | 05/21/2025    | 13:29         | LB135868   |
|           | Chromium | 10.3   | 10.0       | 103        | 80 - 120               | P | 05/21/2025    | 13:29         | LB135868   |
|           | Copper   | 21.8   | 20.0       | 109        | 80 - 120               | P | 05/21/2025    | 13:29         | LB135868   |
|           | Iron     | 105    | 100        | 105        | 80 - 120               | P | 05/21/2025    | 13:29         | LB135868   |
|           | Lead     | 11.0   | 12.0       | 92         | 80 - 120               | P | 05/21/2025    | 13:29         | LB135868   |
|           | Selenium | 19.5   | 20.0       | 98         | 80 - 120               | P | 05/21/2025    | 13:29         | LB135868   |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Environmental, Inc **SDG No.:** Q2090

**Contract:** TULL01 **Lab Code:** CHEM **Case No.:** Q2090 **SAS No.:** Q2090

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Arsenic  | 4850           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 14:09            | LB135868      |
|           | Cadmium  | 2430           | 2500       | 97            | 90 - 110                  | P | 05/21/2025       | 14:09            | LB135868      |
|           | Chromium | 984            | 1000       | 98            | 90 - 110                  | P | 05/21/2025       | 14:09            | LB135868      |
|           | Copper   | 1230           | 1250       | 98            | 90 - 110                  | P | 05/21/2025       | 14:09            | LB135868      |
|           | Iron     | 5060           | 5000       | 101           | 90 - 110                  | P | 05/21/2025       | 14:09            | LB135868      |
|           | Lead     | 4900           | 5000       | 98            | 90 - 110                  | P | 05/21/2025       | 14:09            | LB135868      |
|           | Selenium | 4860           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 14:09            | LB135868      |
| CCV02     | Arsenic  | 4860           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 15:00            | LB135868      |
|           | Cadmium  | 2460           | 2500       | 98            | 90 - 110                  | P | 05/21/2025       | 15:00            | LB135868      |
|           | Chromium | 995            | 1000       | 100           | 90 - 110                  | P | 05/21/2025       | 15:00            | LB135868      |
|           | Copper   | 1220           | 1250       | 98            | 90 - 110                  | P | 05/21/2025       | 15:00            | LB135868      |
|           | Iron     | 4930           | 5000       | 99            | 90 - 110                  | P | 05/21/2025       | 15:00            | LB135868      |
|           | Lead     | 4910           | 5000       | 98            | 90 - 110                  | P | 05/21/2025       | 15:00            | LB135868      |
|           | Selenium | 4860           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 15:00            | LB135868      |
| CCV03     | Arsenic  | 4780           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 16:02            | LB135868      |
|           | Cadmium  | 2390           | 2500       | 96            | 90 - 110                  | P | 05/21/2025       | 16:02            | LB135868      |
|           | Chromium | 983            | 1000       | 98            | 90 - 110                  | P | 05/21/2025       | 16:02            | LB135868      |
|           | Copper   | 1210           | 1250       | 97            | 90 - 110                  | P | 05/21/2025       | 16:02            | LB135868      |
|           | Iron     | 4860           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 16:02            | LB135868      |
|           | Lead     | 4810           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 16:02            | LB135868      |
|           | Selenium | 4790           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 16:02            | LB135868      |
| CCV04     | Arsenic  | 4680           | 5000       | 94            | 90 - 110                  | P | 05/21/2025       | 17:27            | LB135868      |
|           | Cadmium  | 2370           | 2500       | 95            | 90 - 110                  | P | 05/21/2025       | 17:27            | LB135868      |
|           | Chromium | 961            | 1000       | 96            | 90 - 110                  | P | 05/21/2025       | 17:27            | LB135868      |
|           | Copper   | 1190           | 1250       | 96            | 90 - 110                  | P | 05/21/2025       | 17:27            | LB135868      |
|           | Iron     | 4790           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 17:27            | LB135868      |
|           | Lead     | 4790           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 17:27            | LB135868      |
|           | Selenium | 4700           | 5000       | 94            | 90 - 110                  | P | 05/21/2025       | 17:27            | LB135868      |
| CCV05     | Arsenic  | 4730           | 5000       | 95            | 90 - 110                  | P | 05/21/2025       | 18:16            | LB135868      |
|           | Cadmium  | 2410           | 2500       | 96            | 90 - 110                  | P | 05/21/2025       | 18:16            | LB135868      |
|           | Chromium | 984            | 1000       | 98            | 90 - 110                  | P | 05/21/2025       | 18:16            | LB135868      |
|           | Copper   | 1200           | 1250       | 96            | 90 - 110                  | P | 05/21/2025       | 18:16            | LB135868      |
|           | Iron     | 4870           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 18:16            | LB135868      |
|           | Lead     | 4830           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 18:16            | LB135868      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Environmental, Inc **SDG No.:** Q2090  
**Contract:** TULL01 **Lab Code:** CHEM **Case No.:** Q2090 **SAS No.:** Q2090  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV05     | Selenium | 4720           | 5000       | 94            | 90 - 110                  | P | 05/21/2025       | 18:16            | LB135868      |
| CCV06     | Arsenic  | 4940           | 5000       | 99            | 90 - 110                  | P | 05/21/2025       | 19:08            | LB135868      |
|           | Cadmium  | 2500           | 2500       | 100           | 90 - 110                  | P | 05/21/2025       | 19:08            | LB135868      |
|           | Chromium | 1010           | 1000       | 101           | 90 - 110                  | P | 05/21/2025       | 19:08            | LB135868      |
|           | Copper   | 1240           | 1250       | 99            | 90 - 110                  | P | 05/21/2025       | 19:08            | LB135868      |
|           | Iron     | 4970           | 5000       | 99            | 90 - 110                  | P | 05/21/2025       | 19:08            | LB135868      |
|           | Lead     | 4990           | 5000       | 100           | 90 - 110                  | P | 05/21/2025       | 19:08            | LB135868      |
|           | Selenium | 4940           | 5000       | 99            | 90 - 110                  | P | 05/21/2025       | 19:08            | LB135868      |
| CCV07     | Arsenic  | 4940           | 5000       | 99            | 90 - 110                  | P | 05/21/2025       | 19:57            | LB135868      |
|           | Cadmium  | 2480           | 2500       | 99            | 90 - 110                  | P | 05/21/2025       | 19:57            | LB135868      |
|           | Chromium | 1000           | 1000       | 100           | 90 - 110                  | P | 05/21/2025       | 19:57            | LB135868      |
|           | Copper   | 1240           | 1250       | 99            | 90 - 110                  | P | 05/21/2025       | 19:57            | LB135868      |
|           | Iron     | 4920           | 5000       | 98            | 90 - 110                  | P | 05/21/2025       | 19:57            | LB135868      |
|           | Lead     | 4950           | 5000       | 99            | 90 - 110                  | P | 05/21/2025       | 19:57            | LB135868      |
|           | Selenium | 4970           | 5000       | 99            | 90 - 110                  | P | 05/21/2025       | 19:57            | LB135868      |
| CCV08     | Arsenic  | 4800           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 20:43            | LB135868      |
|           | Cadmium  | 2410           | 2500       | 96            | 90 - 110                  | P | 05/21/2025       | 20:43            | LB135868      |
|           | Chromium | 974            | 1000       | 97            | 90 - 110                  | P | 05/21/2025       | 20:43            | LB135868      |
|           | Copper   | 1210           | 1250       | 97            | 90 - 110                  | P | 05/21/2025       | 20:43            | LB135868      |
|           | Iron     | 4720           | 5000       | 94            | 90 - 110                  | P | 05/21/2025       | 20:43            | LB135868      |
|           | Lead     | 4820           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 20:43            | LB135868      |
|           | Selenium | 4820           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 20:43            | LB135868      |
| CCV09     | Arsenic  | 4790           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 21:29            | LB135868      |
|           | Cadmium  | 2440           | 2500       | 98            | 90 - 110                  | P | 05/21/2025       | 21:29            | LB135868      |
|           | Chromium | 993            | 1000       | 99            | 90 - 110                  | P | 05/21/2025       | 21:29            | LB135868      |
|           | Copper   | 1200           | 1250       | 96            | 90 - 110                  | P | 05/21/2025       | 21:29            | LB135868      |
|           | Iron     | 4750           | 5000       | 95            | 90 - 110                  | P | 05/21/2025       | 21:29            | LB135868      |
|           | Lead     | 4890           | 5000       | 98            | 90 - 110                  | P | 05/21/2025       | 21:29            | LB135868      |
|           | Selenium | 4790           | 5000       | 96            | 90 - 110                  | P | 05/21/2025       | 21:29            | LB135868      |
| CCV10     | Arsenic  | 4860           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 21:50            | LB135868      |
|           | Cadmium  | 2430           | 2500       | 97            | 90 - 110                  | P | 05/21/2025       | 21:50            | LB135868      |
|           | Chromium | 987            | 1000       | 99            | 90 - 110                  | P | 05/21/2025       | 21:50            | LB135868      |
|           | Copper   | 1230           | 1250       | 98            | 90 - 110                  | P | 05/21/2025       | 21:50            | LB135868      |
|           | Iron     | 4730           | 5000       | 95            | 90 - 110                  | P | 05/21/2025       | 21:50            | LB135868      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Environmental, Inc **SDG No.:** Q2090  
**Contract:** TULL01 **Lab Code:** CHEM **Case No.:** Q2090 **SAS No.:** Q2090  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV10     | Lead     | 4870           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 21:50            | LB135868      |
|           | Selenium | 4860           | 5000       | 97            | 90 - 110                  | P | 05/21/2025       | 21:50            | LB135868      |