



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q2090
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2090 **SAS No.:** Q2090
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	17.9	20.0	89	65 - 135	P	05/21/2025	13:43	LB135868
	Cadmium	6.27	6.0	104	65 - 135	P	05/21/2025	13:43	LB135868
	Chromium	10.6	10.0	106	65 - 135	P	05/21/2025	13:43	LB135868
	Copper	22.4	20.0	112	65 - 135	P	05/21/2025	13:43	LB135868
	Iron	106	100	106	65 - 135	P	05/21/2025	13:43	LB135868
	Lead	11.1	12.0	92	65 - 135	P	05/21/2025	13:43	LB135868
	Selenium	18.3	20.0	92	65 - 135	P	05/21/2025	13:43	LB135868