



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2090

Contract: TULL01 Lab Code: CHEM

Case No.: Q2090

SAS No.: Q2090

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	13:36	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	13:36	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	13:36	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	13:36	LB135868
	Iron	26.9	+/-50	J	100	P	05/21/2025	13:36	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	13:36	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	13:36	LB135868

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc		SDG No.: Q2090							
Contract: TULL01	Lab Code: CHEM	Case No.: Q2090	SAS No.: Q2090						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	14:13	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	14:13	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	14:13	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	14:13	LB135868
	Iron	23.4	+/-50	U	100	P	05/21/2025	14:13	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	14:13	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	14:13	LB135868
CCB02	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	15:04	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	15:04	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	15:04	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	15:04	LB135868
	Iron	23.4	+/-50	U	100	P	05/21/2025	15:04	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	15:04	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	15:04	LB135868
CCB03	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	16:25	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	16:25	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	16:25	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	16:25	LB135868
	Iron	23.4	+/-50	U	100	P	05/21/2025	16:25	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	16:25	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	16:25	LB135868
CCB04	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	17:32	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	17:32	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	17:32	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	17:32	LB135868
	Iron	23.4	+/-50	U	100	P	05/21/2025	17:32	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	17:32	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	17:32	LB135868
CCB05	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	18:20	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	18:20	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	18:20	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	18:20	LB135868
	Iron	23.4	+/-50	U	100	P	05/21/2025	18:20	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	18:20	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	18:20	LB135868
CCB06	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	19:16	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	19:16	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	19:16	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	19:16	LB135868

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc		SDG No.: Q2090							
Contract: TULL01	Lab Code: CHEM	Case No.: Q2090	SAS No.: Q2090						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Iron	23.4	+/-50	U	100	P	05/21/2025	19:16	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	19:16	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	19:16	LB135868
CCB07	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	20:02	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	20:02	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	20:02	LB135868
CCB08	Copper	4.60	+/-10	U	20.0	P	05/21/2025	20:02	LB135868
	Iron	23.4	+/-50	U	100	P	05/21/2025	20:02	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	20:02	LB135868
CCB09	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	20:02	LB135868
	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	20:48	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	20:48	LB135868
CCB10	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	20:48	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	20:48	LB135868
	Iron	23.4	+/-50	U	100	P	05/21/2025	20:48	LB135868
CCB09	Lead	2.30	+/-6	U	12.0	P	05/21/2025	20:48	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	20:48	LB135868
	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	21:34	LB135868
CCB10	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	21:34	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	21:34	LB135868
	Copper	4.60	+/-10	U	20.0	P	05/21/2025	21:34	LB135868
CCB09	Iron	23.4	+/-50	U	100	P	05/21/2025	21:34	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	21:34	LB135868
	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	21:34	LB135868
CCB10	Arsenic	5.12	+/-10	U	20.0	P	05/21/2025	21:54	LB135868
	Cadmium	0.50	+/-3	U	6.00	P	05/21/2025	21:54	LB135868
	Chromium	2.12	+/-5	U	10.0	P	05/21/2025	21:54	LB135868
CCB09	Copper	4.60	+/-10	U	20.0	P	05/21/2025	21:54	LB135868
	Iron	23.4	+/-50	U	100	P	05/21/2025	21:54	LB135868
	Lead	2.30	+/-6	U	12.0	P	05/21/2025	21:54	LB135868
CCB10	Selenium	9.64	+/-10	U	20.0	P	05/21/2025	21:54	LB135868