

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q2090
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2090 **SAS No.:** Q2090
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.10			-20	20	05/21/2025	13:52	LB135868
	Cadmium	2.19	1.0	219	-5	7	05/21/2025	13:52	LB135868
	Chromium	55.3	52.0	106	42	62	05/21/2025	13:52	LB135868
	Copper	3.11	2.0	156	-18	22	05/21/2025	13:52	LB135868
	Iron	91300	101000	90	85600	116500	05/21/2025	13:52	LB135868
	Lead	4.38			-12	12	05/21/2025	13:52	LB135868
	Selenium	-13.8			-20	20	05/21/2025	13:52	LB135868
ICSAB01	Arsenic	108	104	104	88.4	120	05/21/2025	13:56	LB135868
	Cadmium	960	972	99	826	1120	05/21/2025	13:56	LB135868
	Chromium	526	542	97	460	624	05/21/2025	13:56	LB135868
	Copper	471	511	92	434	588	05/21/2025	13:56	LB135868
	Iron	91200	99300	92	84400	114500	05/21/2025	13:56	LB135868
	Lead	48.7	49.0	99	37	61	05/21/2025	13:56	LB135868
	Selenium	33.9	46.0	74	26	66	05/21/2025	13:56	LB135868
ICSA	Arsenic	26.9			-20	20	05/21/2025	14:00	LB135868
	Cadmium	-1.84	1.0	184	-5	7	05/21/2025	14:00	LB135868
	Chromium	58.8	52.0	113	42	62	05/21/2025	14:00	LB135868
	Copper	11.6	2.0	580	-18	22	05/21/2025	14:00	LB135868
	Iron	89000	101000	88	85600	116500	05/21/2025	14:00	LB135868
	Lead	-3.72			-12	12	05/21/2025	14:00	LB135868
	Selenium	62.3			-20	20	05/21/2025	14:00	LB135868
ICSAB	Arsenic	33.2	104	32	88.4	120	05/21/2025	14:05	LB135868
	Cadmium	968	972	100	826	1120	05/21/2025	14:05	LB135868
	Chromium	549	542	101	460	624	05/21/2025	14:05	LB135868
	Copper	494	511	97	434	588	05/21/2025	14:05	LB135868
	Iron	95000	99300	96	84400	114500	05/21/2025	14:05	LB135868
	Lead	23.0	49.0	47	37	61	05/21/2025	14:05	LB135868
	Selenium	12.8	46.0	28	26	66	05/21/2025	14:05	LB135868