

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q2090
contract: TULL01 **lab code:** CHEM **case no.:** Q2090 **sas no.:** Q2090
matrix: Water **sample id:** Q2090-02 **client id:** 002-35TH-AVE(APR)MS
Percent Solids for Sample: NA **Spiked ID:** Q2090-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	465		10.0	U	400	116		P
Cadmium	ug/L	75 - 125	115		3.00	U	100	115		P
Chromium	ug/L	75 - 125	237		4.94	J	200	116		P
Copper	ug/L	75 - 125	187		17.8		150	113		P
Iron	ug/L	75 - 125	9800		8070		1500	115		P
Lead	ug/L	75 - 125	567		12.4		500	111		P
Selenium	ug/L	75 - 125	1030		10.0	U	1000	103		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q2090
contract: TULL01 **lab code:** CHEM **case no.:** Q2090 **sas no.:** Q2090
matrix: Water **sample id:** Q2090-02 **client id:** 002-35TH-AVE(APR)MSD
Percent Solids for Sample: NA **Spiked ID:** Q2090-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	434		10.0	U	400	108		P
Cadmium	ug/L	75 - 125	106		3.00	U	100	106		P
Chromium	ug/L	75 - 125	219		4.94	J	200	107		P
Copper	ug/L	75 - 125	173		17.8		150	103		P
Iron	ug/L	75 - 125	8340		8070		1500	18		P
Lead	ug/L	75 - 125	523		12.4		500	102		P
Selenium	ug/L	75 - 125	958		10.0	U	1000	96		P