

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2090

Contract: TULL01

Lab Code: CHEM

Case No.: Q2090

SAS No.: Q2090

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168107BS							
Arsenic	ug/L	400	379		95	85 - 115	P
Cadmium	ug/L	100	96.0		96	85 - 115	P
Chromium	ug/L	200	190		95	85 - 115	P
Copper	ug/L	150	145		97	85 - 115	P
Iron	ug/L	1500	1380		92	85 - 115	P
Lead	ug/L	500	468		94	85 - 115	P
Selenium	ug/L	1000	962		96	85 - 115	P