

Laboratory Control Sample Summary

Client: Willets Point Asphalt Corporation

SDG No.: Q2091

Project: SPDES 2025

Run No.: LB135861

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135861BS							
TSS	mg/L	550	533		97	1	90-110	05/21/2025