

## Report of Analysis

Client:	PSEG	Date Collected:	05/20/25 09:58
Project:	Bergen Point Substation	Date Received:	05/20/25
Client Sample ID:	SOIL-DRUM	SDG No.:	Q2093
Lab Sample ID:	Q2093-02	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	8.39	H	1	0	0	pH		05/20/25 17:25	9045D
Ignitability	NO		1	0	0	oC		05/21/25 10:52	1030
Reactive Cyanide	0.014	J	1	0.0083	0.050	mg/Kg	05/23/25 09:00	05/23/25 12:04	9012B
Reactive Sulfide	4.77	J	1	0.20	10.0	mg/Kg	05/22/25 14:00	05/22/25 16:21	9034

Comments: pH result reported at temperature 20.6 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits