

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** Q2098
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q2098 **SAS No.:** Q2098
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	523	510	102	95 - 105	P	05/29/2025	11:55	LB135952
	Zinc	1010	1000	101	95 - 105	P	05/29/2025	11:55	LB135952

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** Q2098
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q2098 **SAS No.:** Q2098
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	20.8	20.0	104	80 - 120	P	05/29/2025	12:15	LB135952
	Zinc	43.6	40.0	109	80 - 120	P	05/29/2025	12:15	LB135952

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** Q2098
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q2098 **SAS No.:** Q2098
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1240	1250	99	90 - 110	P	05/29/2025	14:02	LB135952
	Zinc	2500	2500	100	90 - 110	P	05/29/2025	14:02	LB135952
CCV02	Copper	1230	1250	98	90 - 110	P	05/29/2025	14:51	LB135952
	Zinc	2510	2500	100	90 - 110	P	05/29/2025	14:51	LB135952
CCV03	Copper	1220	1250	98	90 - 110	P	05/29/2025	15:39	LB135952
	Zinc	2450	2500	98	90 - 110	P	05/29/2025	15:39	LB135952
CCV04	Copper	1220	1250	97	90 - 110	P	05/29/2025	16:24	LB135952
	Zinc	2370	2500	95	90 - 110	P	05/29/2025	16:24	LB135952
CCV05	Copper	1230	1250	99	90 - 110	P	05/29/2025	17:10	LB135952
	Zinc	2470	2500	99	90 - 110	P	05/29/2025	17:10	LB135952
CCV06	Copper	1240	1250	99	90 - 110	P	05/29/2025	17:57	LB135952
	Zinc	2490	2500	100	90 - 110	P	05/29/2025	17:57	LB135952
CCV07	Copper	1180	1250	95	90 - 110	P	05/29/2025	18:35	LB135952
	Zinc	2870	2500	115	90 - 110	P	05/29/2025	18:35	LB135952