



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tris Pharma, Inc. **SDG No.:** Q2098
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q2098 **SAS No.:** Q2098
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	21.0	20.0	105	65 - 135	P	05/29/2025	12:23	LB135952
	Zinc	42.7	40.0	107	65 - 135	P	05/29/2025	12:23	LB135952