

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tris Pharma, Inc. **SDG No.:** Q2098
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q2098 **SAS No.:** Q2098
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	2.75	2.0	138	-18	22	05/29/2025	12:41	LB135952
	Zinc	11.7			-40	40	05/29/2025	12:41	LB135952
ICSAB01	Copper	515	511	101	434	588	05/29/2025	12:46	LB135952
	Zinc	1080	952	113	809	1095	05/29/2025	12:46	LB135952
ICSA	Copper	9.71	2.0	486	-18	22	05/29/2025	12:52	LB135952
	Zinc	-58.5			-40	40	05/29/2025	12:52	LB135952
ICSAB	Copper	528	511	103	434	588	05/29/2025	12:56	LB135952
	Zinc	962	952	101	809	1095	05/29/2025	12:56	LB135952