

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tris Pharma, Inc. **level:** low **sdg no.:** Q2098
contract: TRIS02 **lab code:** CHEM **case no.:** Q2098 **sas no.:** Q2098
matrix: Water **sample id:** Q2098-04 **client id:** OUTFALL-DSN-002MS
Percent Solids for Sample: NA **Spiked ID:** Q2098-04MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	228		72.6		150	103		P
Zinc	ug/L	75 - 125	149		45.8		100	103		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tris Pharma, Inc.</u>	level: <u>low</u>	sdg no.: <u>Q2098</u>
contract: <u>TRIS02</u>	lab code: <u>CHEM</u>	case no.: <u>Q2098</u> sas no.: <u>Q2098</u>
matrix: <u>Water</u>	sample id: <u>Q2098-04</u>	client id: <u>OUTFALL-DSN-002MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2098-04MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	222		72.6		150	100		P
Zinc	ug/L	75 - 125	146		45.8		100	101		P