

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tris Pharma, Inc. **SDG No.:** Q2098
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q2098 **SAS No.:** Q2098

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168165BS							
Copper	ug/L	150	147		98	85 - 115	P
Zinc	ug/L	100	95.7		96	85 - 115	P