

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TRIS02

Lab Code: CHEM Case No.: Q2098 SAS No.: Q2098 SDG No.: Q2098

Instrument ID: MSVOA_N Calibration Date/Time: 05/29/2025 11:20

Lab File ID: VN086807.D Init. Calib. Date(s): 05/29/2025 05/29/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:56 12:32

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Ethyl Acetate	0.395	0.415		5.06	
Isopropyl Acetate	0.678	0.694		2.36	
Acetone	0.232	0.235		1.29	
Methylene Chloride	2.055	2.075		0.97	
1,2-Dichloroethane-d4	2.061	2.082		1.02	
Toluene-d8	1.377	1.392		1.09	
4-Bromofluorobenzene	0.471	0.468		-0.64	
n-amyl Acetate	0.545	0.559		2.57	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.