

Report of Analysis

Client:	Tris Pharma, Inc.		Date Collected:	05/20/25	
Project:	Quarterly		Date Received:	05/20/25	
Client Sample ID:	OUTFALL-DSN-001MS		SDG No.:	Q2098	
Lab Sample ID:	Q2098-02MS		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086820.D	1		05/29/25 17:25	VN052925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
141-78-6	Ethyl Acetate	27.8		1.30	5.00	ug/L
108-21-4	Isopropyl Acetate	29.0		1.00	5.00	ug/L
67-64-1	Acetone	410		4.60	25.0	ug/L
75-09-2	Methylene Chloride	30.0		0.86	5.00	ug/L
628-63-7	n-amyl Acetate	32.9		0.81	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	31.1		91 - 110	104%	SPK: 30
2037-26-5	Toluene-d8	30.2		91 - 112	101%	SPK: 30
460-00-4	4-Bromofluorobenzene	30.0		63 - 112	100%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	28700	7.812			
540-36-3	1,4-Difluorobenzene	167000	9.1			
3114-55-4	Chlorobenzene-d5	145000	11.859			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products