

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN052925

Review By	John Carlone	Review On	5/30/2025 9:55:50 AM
Supervise By	Mahesh Dadoda	Supervise On	5/30/2025 12:24:58 PM
SubDirectory	VN052925	HP Acquire Method	HP Processing Method 624N052925W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134047
Initial Calibration Stds	VP134060,VP134061,VP134062,VP134063,VP134064
CCC	VP134048,VP134066,VP134067
Internal Standard/PEM	
ICV/I.BLK	VP134065
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086803.D	29 May 2025 08:57		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN086804.D	29 May 2025 09:56	Need ICAL	JC\MD	Not Ok
3	IBLK	IBLK	VN086805.D	29 May 2025 10:32		JC\MD	Ok
4	VSTDICC005	VSTDICC005	VN086806.D	29 May 2025 10:56		JC\MD	Ok,M
5	VSTDICCC020	VSTDICCC020	VN086807.D	29 May 2025 11:20		JC\MD	Ok,M
6	VSTDICC050	VSTDICC050	VN086808.D	29 May 2025 11:44		JC\MD	Ok,M
7	VSTDICC100	VSTDICC100	VN086809.D	29 May 2025 12:08		JC\MD	Ok,M
8	VSTDICC150	VSTDICC150	VN086810.D	29 May 2025 12:32		JC\MD	Ok,M
9	IBLK	IBLK	VN086811.D	29 May 2025 12:56		JC\MD	Ok
10	VSTDICV020	ICVVN052925	VN086812.D	29 May 2025 13:29		JC\MD	Ok,M
11	VN0529WBS01	VN0529WBS01	VN086813.D	29 May 2025 14:04		JC\MD	Ok,M
12	VN0529WBSD01	VN0529WBSD01	VN086814.D	29 May 2025 14:28		JC\MD	Ok,M
13	VN0529WBL01	VN0529WBL01	VN086815.D	29 May 2025 14:55		JC\MD	Ok
14	Q2126-07 2.5PPB	LOD-MDL-WATER-01-	VN086816.D	29 May 2025 15:29	End CCC Failed	JC\MD	ReRun
15	Q2126-08 5.0PPB	LOQ-WATER-02-QT2-2	VN086817.D	29 May 2025 16:02	End CCC Failed	JC\MD	ReRun
16	Q2098-04	OUTFALL-DSN-002	VN086818.D	29 May 2025 16:37	vial A pH<2 Surrogate Fail;Internal Standard Fail;	JC\MD	Ok,M
17	Q2098-01	OUTFALL-DSN-001	VN086819.D	29 May 2025 17:01	vial A pH<2	JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN052925

Review By	John Carlone	Review On	5/30/2025 9:55:50 AM
Supervise By	Maresh Dadoda	Supervise On	5/30/2025 12:24:58 PM
SubDirectory	VN052925	HP Acquire Method	HP Processing Method 624N052925W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134047		
Initial Calibration Stds	VP134060,VP134061,VP134062,VP134063,VP134064		
CCC	VP134048,VP134066,VP134067		
Internal Standard/PEM			
ICV/I.BLK	VP134065		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2098-02MS	OUTFALL-DSN-001MS	VN086820.D	29 May 2025 17:25	Recovery failed above 50% compound of client ask parameter	JC\MD	Ok,M
19	Q2098-03MSD	OUTFALL-DSN-001MS	VN086821.D	29 May 2025 17:49		JC\MD	Ok,M
20	IBLK	IBLK	VN086822.D	29 May 2025 18:13		JC\MD	Ok
21	Q2098-04RE	OUTFALL-DSN-002RE	VN086823.D	29 May 2025 18:37	vial B pH<2 Not purged	JC\MD	Not Ok
22	VSTDCCC020	VSTDCCC020EC	VN086824.D	29 May 2025 19:01	END CCC fail for comp.#61	JC\MD	Ok,M

M : Manual Integration