



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Metals

- 3a -

#### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| Client:   | CDM Smith |                |                     |              | SDG No.:  | Q2100 |                  |                  |               |  |
|-----------|-----------|----------------|---------------------|--------------|-----------|-------|------------------|------------------|---------------|--|
| Contract: | CAMP02    | Lab Code:      | CHEM                |              | Case No.: | Q2100 | SAS No.:         | Q2100            |               |  |
| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL      | M     | Analysis<br>Date | Analysis<br>Time | Run<br>Number |  |
| ICB16     | Mercury   | 0.20           | +/-0.20             | U            | 0.20      | CV    | 05/23/2025       | 12:31            | LB135901      |  |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> CDM Smith |         | <b>SDG No.:</b> Q2100 |                     |                        |      |                       |                  |                  |               |
|--------------------------|---------|-----------------------|---------------------|------------------------|------|-----------------------|------------------|------------------|---------------|
| <b>Contract:</b> CAMP02  |         | <b>Lab Code:</b> CHEM |                     | <b>Case No.:</b> Q2100 |      | <b>SAS No.:</b> Q2100 |                  |                  |               |
| Sample ID                | Analyte | Result<br>ug/L        | Acceptance<br>Limit | Conc<br>Qual           | CRQL | M                     | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB60                    | Mercury | 0.20                  | +/-0.20             | U                      | 0.20 | CV                    | 05/23/2025       | 12:36            | LB135901      |
| CCB61                    | Mercury | 0.20                  | +/-0.20             | U                      | 0.20 | CV                    | 05/23/2025       | 13:03            | LB135901      |
| CCB62                    | Mercury | 0.20                  | +/-0.20             | U                      | 0.20 | CV                    | 05/23/2025       | 13:31            | LB135901      |
| CCB63                    | Mercury | 0.20                  | +/-0.20             | U                      | 0.20 | CV                    | 05/23/2025       | 13:58            | LB135901      |
| CCB64                    | Mercury | 0.20                  | +/-0.20             | U                      | 0.20 | CV                    | 05/23/2025       | 14:23            | LB135901      |
| CCB65                    | Mercury | 0.20                  | +/-0.20             | U                      | 0.20 | CV                    | 05/23/2025       | 14:39            | LB135901      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> CDM Smith |                       | <b>SDG No.:</b> Q2100  |                       |              |      |   |                  |                  |               |
|--------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> CAMP02  | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q2100 | <b>SAS No.:</b> Q2100 |              |      |   |                  |                  |               |
| Sample ID                | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| ICB01                    | Aluminum              | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Barium                | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Iron                  | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 15:26            | LB135969      |
|                          | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 15:26            | LB135969      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> CDM Smith |                       | <b>SDG No.:</b> Q2100  |                       |              |      |   |                  |                  |               |
|--------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> CAMP02  | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q2100 | <b>SAS No.:</b> Q2100 |              |      |   |                  |                  |               |
| Sample ID                | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB01                    | Aluminum              | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Barium                | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Iron                  | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 16:06            | LB135969      |
|                          | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 16:06            | LB135969      |
| CCB02                    | Aluminum              | 11.8                   | +/-100                | J            | 100  | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Barium                | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Iron                  | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 16:56            | LB135969      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> CDM Smith |                       | <b>SDG No.:</b> Q2100  |                       |              |      |   |                  |                  |               |
|--------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> CAMP02  | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q2100 | <b>SAS No.:</b> Q2100 |              |      |   |                  |                  |               |
| Sample ID                | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB02                    | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 16:56            | LB135969      |
|                          | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 16:56            | LB135969      |
| CCB03                    | Aluminum              | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Barium                | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Iron                  | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 17:45            | LB135969      |
|                          | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 17:45            | LB135969      |
| CCB04                    | Aluminum              | 13.3                   | +/-100                | J            | 100  | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Barium                | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Iron                  | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 05/30/2025       | 18:32            | LB135969      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

| <b>Client:</b> CDM Smith |                       | <b>SDG No.:</b> Q2100  |                       |              |      |   |                  |                  |               |
|--------------------------|-----------------------|------------------------|-----------------------|--------------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> CAMP02  | <b>Lab Code:</b> CHEM | <b>Case No.:</b> Q2100 | <b>SAS No.:</b> Q2100 |              |      |   |                  |                  |               |
| Sample ID                | Analyte               | Result<br>ug/L         | Acceptance<br>Limit   | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB04                    | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 18:32            | LB135969      |
|                          | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 18:32            | LB135969      |
| CCB05                    | Aluminum              | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Barium                | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Chromium              | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Cobalt                | 30.0                   | +/-30.0               | U            | 30.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Copper                | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Iron                  | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Lead                  | 12.0                   | +/-12.0               | U            | 12.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Magnesium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Manganese             | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Nickel                | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Potassium             | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Selenium              | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Silver                | 10.0                   | +/-10.0               | U            | 10.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Sodium                | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Thallium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Vanadium              | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 19:02            | LB135969      |
|                          | Zinc                  | 40.0                   | +/-40.0               | U            | 40.0 | P | 05/30/2025       | 19:02            | LB135969      |
| CCB06                    | Aluminum              | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 19:34            | LB135969      |
|                          | Antimony              | 50.0                   | +/-50.0               | U            | 50.0 | P | 05/30/2025       | 19:34            | LB135969      |
|                          | Arsenic               | 20.0                   | +/-20.0               | U            | 20.0 | P | 05/30/2025       | 19:34            | LB135969      |
|                          | Barium                | 100                    | +/-100                | U            | 100  | P | 05/30/2025       | 19:34            | LB135969      |
|                          | Beryllium             | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 19:34            | LB135969      |
|                          | Cadmium               | 6.00                   | +/-6.00               | U            | 6.00 | P | 05/30/2025       | 19:34            | LB135969      |
|                          | Calcium               | 2000                   | +/-2000               | U            | 2000 | P | 05/30/2025       | 19:34            | LB135969      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> CDM Smith |           | <b>SDG No.:</b> Q2100 |                     |                        |      |                       |                  |                  |               |
|--------------------------|-----------|-----------------------|---------------------|------------------------|------|-----------------------|------------------|------------------|---------------|
| <b>Contract:</b> CAMP02  |           | <b>Lab Code:</b> CHEM |                     | <b>Case No.:</b> Q2100 |      | <b>SAS No.:</b> Q2100 |                  |                  |               |
| Sample ID                | Analyte   | Result<br>ug/L        | Acceptance<br>Limit | Conc<br>Qual           | CRQL | M                     | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB06                    | Chromium  | 10.0                  | +/-10.0             | U                      | 10.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Cobalt    | 30.0                  | +/-30.0             | U                      | 30.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Copper    | 20.0                  | +/-20.0             | U                      | 20.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Iron      | 100                   | +/-100              | U                      | 100  | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Lead      | 12.0                  | +/-12.0             | U                      | 12.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Magnesium | 2000                  | +/-2000             | U                      | 2000 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Manganese | 20.0                  | +/-20.0             | U                      | 20.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Nickel    | 40.0                  | +/-40.0             | U                      | 40.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Potassium | 2000                  | +/-2000             | U                      | 2000 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Selenium  | 20.0                  | +/-20.0             | U                      | 20.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Silver    | 10.0                  | +/-10.0             | U                      | 10.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Sodium    | 2000                  | +/-2000             | U                      | 2000 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Thallium  | 40.0                  | +/-40.0             | U                      | 40.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Vanadium  | 40.0                  | +/-40.0             | U                      | 40.0 | P                     | 05/30/2025       | 19:34            | LB135969      |
|                          | Zinc      | 40.0                  | +/-40.0             | U                      | 40.0 | P                     | 05/30/2025       | 19:34            | LB135969      |