

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2100
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2100 **SAS No.:** Q2100

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168117BS							
Aluminum	mg/Kg	100	96.9		97	80 - 120	P
Antimony	mg/Kg	40.0	42.8		107	80 - 120	P
Arsenic	mg/Kg	40.0	41.9		105	80 - 120	P
Barium	mg/Kg	10.0	9.64		96	80 - 120	P
Beryllium	mg/Kg	10.0	9.71		97	80 - 120	P
Cadmium	mg/Kg	10.0	10.3		103	80 - 120	P
Calcium	mg/Kg	50.0	50.6	J	101	80 - 120	P
Chromium	mg/Kg	20.0	20.9		104	80 - 120	P
Cobalt	mg/Kg	10.0	10.2		102	80 - 120	P
Copper	mg/Kg	15.0	16.2		108	80 - 120	P
Iron	mg/Kg	150	163		109	80 - 120	P
Lead	mg/Kg	50.0	50.1		100	80 - 120	P
Magnesium	mg/Kg	100	95.4	J	95	80 - 120	P
Manganese	mg/Kg	10.0	9.88		99	80 - 120	P
Nickel	mg/Kg	25.0	25.9		104	80 - 120	P
Potassium	mg/Kg	500	508		102	80 - 120	P
Selenium	mg/Kg	100	107		107	80 - 120	P
Silver	mg/Kg	3.8	3.80		100	80 - 120	P
Sodium	mg/Kg	150	149		99	80 - 120	P
Thallium	mg/Kg	100	105		105	80 - 120	P
Vanadium	mg/Kg	15.0	14.7		98	80 - 120	P
Zinc	mg/Kg	10.0	9.97		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2100
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2100 **SAS No.:** Q2100

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168146BS Mercury	mg/Kg	0.26	0.25		94	80 - 120	CV