

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168137BS

Contract: CAMP02

Lab Code: CHEM **Case No.:** Q2100

SAS No.: Q2100 **SDG NO.:** Q2100

Lab Sample ID: PB168137BS

Date(s) Analyzed: 05/23/2025 05/23/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP **ID:** 0.32 (mm)

GC Column:(2): RTX-CLP2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	7.78	7.73	7.83	164	6.5
	2	8.32	8.27	8.37	175	
2,4-D	1	8.00	7.95	8.05	183	6.8
	2	8.63	8.58	8.68	171	
2,4,5-TP(Silvex)	1	8.84	8.79	8.89	167	0.6
	2	9.51	9.46	9.56	168	
2,4,5-T	1	9.13	9.08	9.18	165	5.6
	2	9.91	9.86	9.96	156	
2,4-DB	1	9.68	9.63	9.73	177	27.7
	2	10.47	10.42	10.52	134	
Dinoseb	1	10.84	10.79	10.89	148	8.5
	2	10.84	10.79	10.89	136	
DICAMBA	1	7.10	7.05	7.15	148	1.3
	2	7.63	7.58	7.68	150	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-21MS

Contract: CAMP02

Lab Code: CHEM **Case No.:** Q2100

SAS No.: Q2100

SDG NO.: Q2100

Lab Sample ID: Q2100-03MS

Date(s) Analyzed: 05/29/2025 05/29/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP **ID:** 0.32 (mm)

GC Column:(2): RTX-CLP2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	7.78	7.73	7.83	123	3.3
	2	8.31	8.26	8.36	119	
2,4-D	1	8.00	7.95	8.05	154	13.9
	2	8.63	8.58	8.68	134	
2,4,5-TP(Silvex)	1	8.84	8.79	8.89	120	2.5
	2	9.50	9.45	9.55	123	
2,4,5-T	1	9.12	9.07	9.17	122	3.3
	2	9.91	9.86	9.96	118	
2,4-DB	1	9.68	9.63	9.73	156	34.6
	2	10.46	10.41	10.51	110	
Dinoseb	1	10.84	10.79	10.89	20.4	1
	2	10.84	10.79	10.89	20.6	
DICAMBA	1	7.10	7.05	7.15	117	7.1
	2	7.62	7.57	7.67	109	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-21MSD

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100

SAS No.: Q2100 SDG NO.: Q2100

Lab Sample ID: Q2100-03MSD

Date(s) Analyzed: 05/29/2025 05/29/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column: (2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	7.78	7.73	7.83	128	12.4
	2	8.32	8.27	8.37	113	
2,4-D	1	8.00	7.95	8.05	146	5.6
	2	8.63	8.58	8.68	138	
2,4,5-TP(Silvex)	1	8.84	8.79	8.89	128	0.8
	2	9.50	9.45	9.55	129	
2,4,5-T	1	9.12	9.07	9.17	127	3.2
	2	9.91	9.86	9.96	123	
2,4-DB	1	9.68	9.63	9.73	158	31.5
	2	10.47	10.42	10.52	115	
Dinoseb	1	10.84	10.79	10.89	20.5	0.5
	2	10.84	10.79	10.89	20.6	
DICAMBA	1	7.10	7.05	7.15	121	5.1
	2	7.62	7.57	7.67	115	