

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/23/2025

Lab Sample No.: HSTDCCC750 Data File : PS030380.D Time Analyzed: 14:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.125	9.033	9.233	713.390	712.500	0.1
2,4,5-TP(Silvex)	8.842	8.750	8.950	723.370	712.500	1.5
2,4-D	8.003	7.910	8.110	698.890	705.000	-0.9
2,4-DB	9.681	9.589	9.789	744.800	712.500	4.5
2,4-DCAA	6.927	6.834	7.034	742.820	750.000	-1.0
DICAMBA	7.103	7.010	7.210	713.160	705.000	1.2
DICHLORPROP	7.783	7.689	7.889	697.750	705.000	-1.0
Dinoseb	10.841	10.751	10.951	680.110	705.000	-3.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/23/2025

Lab Sample No.: HSTDCCC750 Data File : PS030380.D Time Analyzed: 14:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.912	9.826	10.026	716.230	712.500	0.5
2,4,5-TP(Silvex)	9.509	9.423	9.623	725.180	712.500	1.8
2,4-D	8.632	8.544	8.744	701.140	705.000	-0.5
2,4-DB	10.470	10.385	10.585	611.550	712.500	-14.2
2,4-DCAA	7.444	7.354	7.554	729.960	750.000	-2.7
DICAMBA	7.629	7.539	7.739	715.670	705.000	1.5
DICHLORPROP	8.320	8.232	8.432	682.800	705.000	-3.1
Dinoseb	10.844	10.758	10.958	691.900	705.000	-1.9