

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/23/2025

Lab Sample No.: HSTDCCC750 Data File : PS030392.D Time Analyzed: 20:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.124	9.033	9.233	747.780	712.500	5.0
2,4,5-TP(Silvex)	8.842	8.750	8.950	749.620	712.500	5.2
2,4-D	8.003	7.910	8.110	727.740	705.000	3.2
2,4-DB	9.682	9.589	9.789	792.870	712.500	11.3
2,4-DCAA	6.927	6.834	7.034	768.250	750.000	2.4
DICAMBA	7.104	7.010	7.210	732.110	705.000	3.8
DICHLORPROP	7.782	7.689	7.889	723.740	705.000	2.7
Dinoseb	10.841	10.751	10.951	725.530	705.000	2.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/23/2025

Lab Sample No.: HSTDCCC750 Data File : PS030392.D Time Analyzed: 20:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.912	9.826	10.026	745.910	712.500	4.7
2,4,5-TP(Silvex)	9.509	9.423	9.623	752.180	712.500	5.6
2,4-D	8.632	8.544	8.744	732.820	705.000	3.9
2,4-DB	10.469	10.385	10.585	645.150	712.500	-9.5
2,4-DCAA	7.444	7.354	7.554	759.110	750.000	1.2
DICAMBA	7.629	7.539	7.739	740.440	705.000	5.0
DICHLORPROP	8.320	8.232	8.432	715.510	705.000	1.5
Dinoseb	10.843	10.758	10.958	722.560	705.000	2.5