

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/24/2025

Lab Sample No.: HSTDCCC750 Data File : PS030404.D Time Analyzed: 02:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.123	9.033	9.233	735.910	712.500	3.3
2,4,5-TP(Silvex)	8.841	8.750	8.950	735.680	712.500	3.3
2,4-D	8.001	7.910	8.110	721.420	705.000	2.3
2,4-DB	9.680	9.589	9.789	785.390	712.500	10.2
2,4-DCAA	6.926	6.834	7.034	761.020	750.000	1.5
DICAMBA	7.102	7.010	7.210	727.020	705.000	3.1
DICHLORPROP	7.782	7.689	7.889	716.650	705.000	1.7
Dinoseb	10.841	10.751	10.951	709.780	705.000	0.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/24/2025

Lab Sample No.: HSTDCCC750 Data File : PS030404.D Time Analyzed: 02:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.911	9.826	10.026	744.760	712.500	4.5
2,4,5-TP(Silvex)	9.508	9.423	9.623	749.740	712.500	5.2
2,4-D	8.631	8.544	8.744	731.630	705.000	3.8
2,4-DB	10.468	10.385	10.585	644.010	712.500	-9.6
2,4-DCAA	7.443	7.354	7.554	759.720	750.000	1.3
DICAMBA	7.628	7.539	7.739	741.390	705.000	5.2
DICHLORPROP	8.319	8.232	8.432	714.060	705.000	1.3
Dinoseb	10.842	10.758	10.958	723.800	705.000	2.7