

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/24/2025

Lab Sample No.: HSTDCCC750 Data File : PS030409.D Time Analyzed: 04:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.124	9.033	9.233	744.740	712.500	4.5
2,4,5-TP(Silvex)	8.841	8.750	8.950	744.430	712.500	4.5
2,4-D	8.002	7.910	8.110	724.030	705.000	2.7
2,4-DB	9.681	9.589	9.789	787.600	712.500	10.5
2,4-DCAA	6.926	6.834	7.034	765.020	750.000	2.0
DICAMBA	7.102	7.010	7.210	728.470	705.000	3.3
DICHLORPROP	7.782	7.689	7.889	719.810	705.000	2.1
Dinoseb	10.840	10.751	10.951	722.330	705.000	2.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/24/2025

Lab Sample No.: HSTDCCC750 Data File : PS030409.D Time Analyzed: 04:30

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.912	9.826	10.026	744.630	712.500	4.5
2,4,5-TP(Silvex)	9.508	9.423	9.623	752.190	712.500	5.6
2,4-D	8.631	8.544	8.744	730.500	705.000	3.6
2,4-DB	10.468	10.385	10.585	648.610	712.500	-9.0
2,4-DCAA	7.443	7.354	7.554	755.690	750.000	0.8
DICAMBA	7.628	7.539	7.739	738.590	705.000	4.8
DICHLORPROP	8.319	8.232	8.432	717.680	705.000	1.8
Dinoseb	10.842	10.758	10.958	726.080	705.000	3.0