

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL05 Date Analyzed: 05/29/2025

Lab Sample No.: HSTDCCC750 Data File : PS030427.D Time Analyzed: 09:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.121	9.033	9.233	707.790	712.500	-0.7
2,4,5-TP(Silvex)	8.839	8.750	8.950	715.020	712.500	0.4
2,4-D	7.999	7.910	8.110	694.600	705.000	-1.5
2,4-DB	9.678	9.589	9.789	753.420	712.500	5.7
2,4-DCAA	6.924	6.834	7.034	731.220	750.000	-2.5
DICAMBA	7.101	7.010	7.210	703.340	705.000	-0.2
DICHLORPROP	7.780	7.689	7.889	685.990	705.000	-2.7
Dinoseb	10.837	10.751	10.951	678.950	705.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL05 Date Analyzed: 05/29/2025

Lab Sample No.: HSTDCCC750 Data File : PS030427.D Time Analyzed: 09:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.908	9.826	10.026	699.550	712.500	-1.8
2,4,5-TP(Silvex)	9.504	9.423	9.623	713.420	712.500	0.1
2,4-D	8.628	8.544	8.744	685.410	705.000	-2.8
2,4-DB	10.465	10.385	10.585	622.520	712.500	-12.6
2,4-DCAA	7.440	7.354	7.554	698.290	750.000	-6.9
DICAMBA	7.625	7.539	7.739	696.340	705.000	-1.2
DICHLORPROP	8.316	8.232	8.432	664.380	705.000	-5.8
Dinoseb	10.837	10.758	10.958	673.270	705.000	-4.5