

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL06 Date Analyzed: 05/29/2025

Lab Sample No.: HSTDCCC750 Data File : PS030434.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.122	9.033	9.233	726.650	712.500	2.0
2,4,5-TP(Silvex)	8.840	8.750	8.950	730.580	712.500	2.5
2,4-D	8.000	7.910	8.110	706.920	705.000	0.3
2,4-DB	9.679	9.589	9.789	771.790	712.500	8.3
2,4-DCAA	6.925	6.834	7.034	742.200	750.000	-1.0
DICAMBA	7.101	7.010	7.210	713.300	705.000	1.2
DICHLORPROP	7.780	7.689	7.889	699.930	705.000	-0.7
Dinoseb	10.839	10.751	10.951	707.390	705.000	0.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL06 Date Analyzed: 05/29/2025

Lab Sample No.: HSTDCCC750 Data File : PS030434.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.908	9.826	10.026	711.790	712.500	-0.1
2,4,5-TP(Silvex)	9.505	9.423	9.623	722.010	712.500	1.3
2,4-D	8.629	8.544	8.744	695.450	705.000	-1.4
2,4-DB	10.466	10.385	10.585	606.820	712.500	-14.8
2,4-DCAA	7.442	7.354	7.554	710.890	750.000	-5.2
DICAMBA	7.626	7.539	7.739	702.680	705.000	-0.3
DICHLORPROP	8.317	8.232	8.432	672.940	705.000	-4.5
Dinoseb	10.839	10.758	10.958	704.100	705.000	-0.1