

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2100</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS030430.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2100-03MSD (Column 1)	TP-21MSD											
	DICAMBA	212.6	0	121	ug/Kg	57		4		10	112	20
	DICHLORPROP	212.6	0	128	ug/Kg	60		3		10	113	20
	2,4-D	212.6	0	146	ug/Kg	69		4		10	144	20
	2,4,5-TP(Silvex)	212.6	0	128	ug/Kg	60		7		10	114	20
	2,4,5-T	212.6	0	127	ug/Kg	60		5		10	115	20
	2,4-DB	212.6	0	158	ug/Kg	74		1		10	140	20
Client Sample ID: Q2100-03MSD (Column 2)	Dinoseb	212.6	0	20.5	ug/Kg	10		0		10	118	20
	TP-21MSD											
	DICAMBA	212.6	0	115	ug/Kg	54		6		10	112	20
	DICHLORPROP	212.6	0	113	ug/Kg	53		6		10	113	20
	2,4-D	212.6	0	138	ug/Kg	65		3		10	144	20
	2,4,5-TP(Silvex)	212.6	0	129	ug/Kg	61		5		10	114	20
	2,4,5-T	212.6	0	123	ug/Kg	58		5		10	115	20
	2,4-DB	212.6	0	115	ug/Kg	54		4		10	140	20
	Dinoseb	212.6	0	20.6	ug/Kg	10		0		10	118	20