

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2100

Analytical Method: 8151A

Client: CDM Smith

Datafile : PS030382.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB168137BS (Column 1)	DICAMBA	166.6	148	ug/Kg	89				72	129
	DICHLORPROP	166.6	164	ug/Kg	98				77	135
	2,4-D	166.6	183	ug/Kg	110				65	144
	2,4,5-TP(Silvex)	166.6	167	ug/Kg	100				74	146
	2,4,5-T	166.6	165	ug/Kg	99				77	134
	2,4-DB	166.6	177	ug/Kg	106				72	122
	Dinoseb	166.6	148	ug/Kg	89				74	132
PB168137BS (Column 2)	DICAMBA	166.6	150	ug/Kg	90				72	129
	DICHLORPROP	166.6	175	ug/Kg	105				77	135
	2,4-D	166.6	171	ug/Kg	103				65	144
	2,4,5-TP(Silvex)	166.6	168	ug/Kg	101				74	146
	2,4,5-T	166.6	156	ug/Kg	94				77	134
	2,4-DB	166.6	134	ug/Kg	80				72	122
	Dinoseb	166.6	136	ug/Kg	82				74	132