

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

Continuing Calib Date: 05/22/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 16:56 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1254-1       | (1) | 6.50       | 6.51      | 6.41                   | 6.61  | 0.01       |
| Aroclor-1254-2       | (2) | 6.72       | 6.72      | 6.62                   | 6.82  | 0.00       |
| Aroclor-1254-3       | (3) | 7.08       | 7.09      | 6.99                   | 7.19  | 0.01       |
| Aroclor-1254-4       | (4) | 7.36       | 7.37      | 7.27                   | 7.47  | 0.01       |
| Aroclor-1254-5       | (5) | 7.78       | 7.79      | 7.69                   | 7.89  | 0.01       |
| Tetrachloro-m-xylene |     | 4.50       | 4.50      | 4.40                   | 4.60  | 0.00       |
| Decachlorobiphenyl   |     | 10.21      | 10.22     | 10.12                  | 10.32 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

Continuing Calib Date: 05/22/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 16:56 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.68       | 5.69      | 5.59      | 5.79 | 0.01       |
| Aroclor-1254-2       | (2) | 5.83       | 5.83      | 5.73      | 5.93 | 0.00       |
| Aroclor-1254-3       | (3) | 6.23       | 6.24      | 6.14      | 6.34 | 0.01       |
| Aroclor-1254-4       | (4) | 6.46       | 6.47      | 6.37      | 6.57 | 0.01       |
| Aroclor-1254-5       | (5) | 6.88       | 6.88      | 6.78      | 6.98 | 0.00       |
| Tetrachloro-m-xylene |     | 3.80       | 3.80      | 3.70      | 3.90 | 0.00       |
| Decachlorobiphenyl   |     | 8.82       | 8.83      | 8.73      | 8.93 | 0.01       |