

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 05/22/2025

Lab Sample No.: AR1660CCC500 Data File : PO111273.D Time Analyzed: 10:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.771	4.673	4.873	523.520	500.000	4.7
Aroclor-1016-2	4.789	4.692	4.892	531.280	500.000	6.3
Aroclor-1016-3	4.846	4.749	4.949	543.870	500.000	8.8
Aroclor-1016-4	4.966	4.869	5.069	535.250	500.000	7.1
Aroclor-1016-5	5.224	5.126	5.326	522.550	500.000	4.5
Aroclor-1260-1	6.263	6.166	6.366	528.000	500.000	5.6
Aroclor-1260-2	6.453	6.355	6.555	525.790	500.000	5.2
Aroclor-1260-3	6.820	6.722	6.922	523.540	500.000	4.7
Aroclor-1260-4	7.080	6.983	7.183	521.080	500.000	4.2
Aroclor-1260-5	7.323	7.225	7.425	525.480	500.000	5.1
Decachlorobiphenyl	8.722	8.624	8.824	50.690	50.000	1.4
Tetrachloro-m-xylene	3.680	3.582	3.782	52.820	50.000	5.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 05/22/2025

Lab Sample No.: AR1660CCC500 Data File : PO111273.D Time Analyzed: 10:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.756	4.658	4.858	501.100	500.000	0.2
Aroclor-1016-2	4.775	4.677	4.877	504.110	500.000	0.8
Aroclor-1016-3	4.950	4.852	5.052	507.060	500.000	1.4
Aroclor-1016-4	4.992	4.894	5.094	504.510	500.000	0.9
Aroclor-1016-5	5.205	5.107	5.307	504.430	500.000	0.9
Aroclor-1260-1	6.236	6.138	6.338	504.970	500.000	1.0
Aroclor-1260-2	6.424	6.326	6.526	502.630	500.000	0.5
Aroclor-1260-3	6.576	6.479	6.679	472.310	500.000	-5.5
Aroclor-1260-4	7.047	6.949	7.149	492.700	500.000	-1.5
Aroclor-1260-5	7.288	7.191	7.391	489.400	500.000	-2.1
Decachlorobiphenyl	8.671	8.574	8.774	48.950	50.000	-2.1
Tetrachloro-m-xylene	3.677	3.579	3.779	51.990	50.000	4.0