

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 05/22/2025

Lab Sample No.: AR1660CCC500 Data File : PO111288.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.768	4.673	4.873	534.780	500.000	7.0
Aroclor-1016-2	4.787	4.692	4.892	538.860	500.000	7.8
Aroclor-1016-3	4.844	4.749	4.949	550.350	500.000	10.1
Aroclor-1016-4	4.964	4.869	5.069	541.370	500.000	8.3
Aroclor-1016-5	5.222	5.126	5.326	525.860	500.000	5.2
Aroclor-1260-1	6.261	6.166	6.366	522.690	500.000	4.5
Aroclor-1260-2	6.450	6.355	6.555	530.360	500.000	6.1
Aroclor-1260-3	6.817	6.722	6.922	509.770	500.000	2.0
Aroclor-1260-4	7.077	6.983	7.183	501.680	500.000	0.3
Aroclor-1260-5	7.321	7.225	7.425	496.330	500.000	-0.7
Decachlorobiphenyl	8.719	8.624	8.824	46.970	50.000	-6.1
Tetrachloro-m-xylene	3.679	3.582	3.782	53.010	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 05/22/2025

Lab Sample No.: AR1660CCC500 Data File : PO111288.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.755	4.658	4.858	509.400	500.000	1.9
Aroclor-1016-2	4.774	4.677	4.877	511.470	500.000	2.3
Aroclor-1016-3	4.949	4.852	5.052	512.030	500.000	2.4
Aroclor-1016-4	4.992	4.894	5.094	503.300	500.000	0.7
Aroclor-1016-5	5.204	5.107	5.307	510.250	500.000	2.1
Aroclor-1260-1	6.235	6.138	6.338	496.920	500.000	-0.6
Aroclor-1260-2	6.423	6.326	6.526	495.270	500.000	-0.9
Aroclor-1260-3	6.575	6.479	6.679	460.070	500.000	-8.0
Aroclor-1260-4	7.046	6.949	7.149	476.950	500.000	-4.6
Aroclor-1260-5	7.288	7.191	7.391	474.510	500.000	-5.1
Decachlorobiphenyl	8.671	8.574	8.774	47.170	50.000	-5.7
Tetrachloro-m-xylene	3.677	3.579	3.779	52.520	50.000	5.0