

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/22/2025

Lab Sample No.: AR1660CCC500 Data File : PP072286.D Time Analyzed: 10:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.654	5.552	5.752	476.140	500.000	-4.8
Aroclor-1016-2	5.676	5.574	5.774	511.140	500.000	2.2
Aroclor-1016-3	5.739	5.636	5.836	497.860	500.000	-0.4
Aroclor-1016-4	5.835	5.734	5.934	515.920	500.000	3.2
Aroclor-1016-5	6.128	6.027	6.227	516.070	500.000	3.2
Aroclor-1260-1	7.246	7.147	7.347	508.170	500.000	1.6
Aroclor-1260-2	7.500	7.401	7.601	504.900	500.000	1.0
Aroclor-1260-3	7.858	7.760	7.960	523.470	500.000	4.7
Aroclor-1260-4	8.082	7.985	8.185	525.410	500.000	5.1
Aroclor-1260-5	8.401	8.303	8.503	527.280	500.000	5.5
Decachlorobiphenyl	10.208	10.113	10.313	53.260	50.000	6.5
Tetrachloro-m-xylene	4.502	4.396	4.596	51.920	50.000	3.8

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/22/2025

Lab Sample No.: AR1660CCC500 Data File : PP072286.D Time Analyzed: 10:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.879	4.777	4.977	517.470	500.000	3.5
Aroclor-1016-2	4.897	4.795	4.995	531.580	500.000	6.3
Aroclor-1016-3	5.074	4.973	5.173	539.340	500.000	7.9
Aroclor-1016-4	5.116	5.015	5.215	544.550	500.000	8.9
Aroclor-1016-5	5.330	5.229	5.429	534.000	500.000	6.8
Aroclor-1260-1	6.364	6.264	6.464	543.290	500.000	8.7
Aroclor-1260-2	6.552	6.453	6.653	513.590	500.000	2.7
Aroclor-1260-3	6.705	6.606	6.806	552.800	500.000	10.6
Aroclor-1260-4	7.176	7.077	7.277	550.020	500.000	10.0
Aroclor-1260-5	7.417	7.319	7.519	543.120	500.000	8.6
Decachlorobiphenyl	8.819	8.723	8.923	54.340	50.000	8.7
Tetrachloro-m-xylene	3.796	3.691	3.891	52.980	50.000	6.0