

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 05/22/2025

Lab Sample No.: AR1660CCC500 Data File : PP072301.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.657	5.552	5.752	488.950	500.000	-2.2
Aroclor-1016-2	5.678	5.574	5.774	518.710	500.000	3.7
Aroclor-1016-3	5.740	5.636	5.836	513.260	500.000	2.7
Aroclor-1016-4	5.838	5.734	5.934	527.750	500.000	5.6
Aroclor-1016-5	6.131	6.027	6.227	526.130	500.000	5.2
Aroclor-1260-1	7.249	7.147	7.347	524.890	500.000	5.0
Aroclor-1260-2	7.502	7.401	7.601	507.290	500.000	1.5
Aroclor-1260-3	7.861	7.760	7.960	530.030	500.000	6.0
Aroclor-1260-4	8.086	7.985	8.185	536.070	500.000	7.2
Aroclor-1260-5	8.405	8.303	8.503	538.410	500.000	7.7
Decachlorobiphenyl	10.213	10.113	10.313	54.040	50.000	8.1
Tetrachloro-m-xylene	4.504	4.396	4.596	53.120	50.000	6.2

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 05/22/2025

Lab Sample No.: AR1660CCC500 Data File : PP072301.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.879	4.777	4.977	528.970	500.000	5.8
Aroclor-1016-2	4.897	4.795	4.995	540.910	500.000	8.2
Aroclor-1016-3	5.075	4.973	5.173	550.960	500.000	10.2
Aroclor-1016-4	5.117	5.015	5.215	552.300	500.000	10.5
Aroclor-1016-5	5.331	5.229	5.429	569.350	500.000	13.9
Aroclor-1260-1	6.364	6.264	6.464	549.050	500.000	9.8
Aroclor-1260-2	6.552	6.453	6.653	518.530	500.000	3.7
Aroclor-1260-3	6.705	6.606	6.806	553.850	500.000	10.8
Aroclor-1260-4	7.176	7.077	7.277	549.370	500.000	9.9
Aroclor-1260-5	7.418	7.319	7.519	549.010	500.000	9.8
Decachlorobiphenyl	8.819	8.723	8.923	57.170	50.000	14.3
Tetrachloro-m-xylene	3.796	3.691	3.891	54.420	50.000	8.8