

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168125BS

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100

SAS No.: Q2100 SDG NO.: Q2100

Lab Sample ID: PB168125BS

Date(s) Analyzed: 05/22/2025 05/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.80	6.75	6.85	17.0	4.2
	2	6.08	6.03	6.13	16.3	
4,4'-DDD	1	6.71	6.66	6.76	20.0	11.6
	2	5.93	5.88	5.98	17.8	
4,4'-DDT	1	7.03	6.98	7.08	13.2	1.5
	2	6.19	6.14	6.24	13.4	
Endrin aldehyde	1	6.92	6.87	6.97	16.4	3.7
	2	6.26	6.21	6.31	15.8	
Endosulfan sulfate	1	7.16	7.11	7.21	16.5	5
	2	6.49	6.44	6.54	15.7	
Methoxychlor	1	7.50	7.45	7.55	13.6	0.7
	2	6.76	6.71	6.81	13.5	
Endrin ketone	1	7.64	7.59	7.69	16.9	5.5
	2	6.99	6.94	7.04	16.0	
alpha-BHC	1	4.01	3.96	4.06	16.9	1.2
	2	3.39	3.34	3.44	16.7	
gamma-BHC (Lindane)	1	4.34	4.29	4.39	17.0	1.8
	2	3.73	3.68	3.78	16.7	
Heptachlor	1	4.94	4.89	4.99	16.8	4.3
	2	4.08	4.03	4.13	16.1	
Aldrin	1	5.28	5.23	5.33	17.3	3.5
	2	4.37	4.32	4.42	16.7	
beta-BHC	1	4.52	4.47	4.57	16.6	0
	2	4.03	3.98	4.08	16.6	
delta-BHC	1	4.77	4.72	4.82	16.9	6.1
	2	4.26	4.21	4.31	15.9	
Heptachlor epoxide	1	5.70	5.65	5.75	17.1	3
	2	4.88	4.83	4.93	16.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168125BS

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100

SAS No.: Q2100

SDG NO.: Q2100

Lab Sample ID: PB168125BS

Date(s) Analyzed: 05/22/2025 05/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.08	6.03	6.13	17.2	13.7
	2	5.25	5.20	5.30	15.0	
gamma-Chlordane	1	5.96	5.91	6.01	17.3	4.1
	2	5.13	5.08	5.18	16.6	
alpha-Chlordane	1	6.04	5.99	6.09	17.2	4.2
	2	5.19	5.14	5.24	16.5	
4,4'-DDE	1	6.21	6.16	6.26	17.1	3.6
	2	5.38	5.33	5.43	16.5	
Dieldrin	1	6.36	6.31	6.41	17.3	4.1
	2	5.52	5.47	5.57	16.6	
Endrin	1	6.59	6.54	6.64	16.8	4.9
	2	5.79	5.74	5.84	16.0	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-21

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

Lab Sample ID: Q2100-03 Date(s) Analyzed: 05/22/2025 05/22/2025

Instrument ID (1): ECD_D Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
alpha-Chlordane	1	6.03	5.98	6.08	0.34	48.8
	2	5.19	5.14	5.24	0.20	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-45MS

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100

SAS No.: Q2100 SDG NO.: Q2100

Lab Sample ID: Q2100-04MS

Date(s) Analyzed: 05/22/2025 05/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	21.3	3.3
	2	6.08	6.03	6.13	20.6	
4,4'-DDD	1	6.71	6.66	6.76	24.3	10.8
	2	5.93	5.88	5.98	21.8	
4,4'-DDT	1	7.02	6.97	7.07	17.0	1.8
	2	6.19	6.14	6.24	16.7	
Endrin aldehyde	1	6.92	6.87	6.97	20.9	4.9
	2	6.26	6.21	6.31	19.9	
Endosulfan sulfate	1	7.15	7.10	7.20	20.8	3.4
	2	6.48	6.43	6.53	20.1	
Methoxychlor	1	7.50	7.45	7.55	16.8	0.6
	2	6.76	6.71	6.81	16.9	
Endrin ketone	1	7.63	7.58	7.68	21.3	5.3
	2	6.99	6.94	7.04	20.2	
alpha-BHC	1	4.00	3.95	4.05	22.6	2.2
	2	3.39	3.34	3.44	22.1	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	22.4	2.7
	2	3.73	3.68	3.78	21.8	
Heptachlor	1	4.93	4.88	4.98	21.7	4.2
	2	4.08	4.03	4.13	20.8	
Aldrin	1	5.27	5.22	5.32	22.5	3.2
	2	4.37	4.32	4.42	21.8	
beta-BHC	1	4.52	4.47	4.57	21.6	2.7
	2	4.03	3.98	4.08	22.2	
delta-BHC	1	4.77	4.72	4.82	24.1	9.6
	2	4.26	4.21	4.31	21.9	
Heptachlor epoxide	1	5.69	5.64	5.74	22.1	3.7
	2	4.87	4.82	4.92	21.3	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-45MS

Contract: CAMP02

Lab Code: CHEM

Case No.: Q2100

SAS No.: Q2100

SDG NO.: Q2100

Lab Sample ID: Q2100-04MS

Date(s) Analyzed: 05/22/2025

05/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.08	6.03	6.13	21.9	3.7
	2	5.25	5.20	5.30	21.1	
gamma-Chlordane	1	5.95	5.90	6.00	22.0	3.2
	2	5.13	5.08	5.18	21.3	
alpha-Chlordane	1	6.03	5.98	6.08	21.9	2.8
	2	5.19	5.14	5.24	21.3	
4,4'-DDE	1	6.20	6.15	6.25	21.9	4.7
	2	5.38	5.33	5.43	20.9	
Dieldrin	1	6.35	6.30	6.40	22.0	4.2
	2	5.51	5.46	5.56	21.1	
Endrin	1	6.58	6.53	6.63	21.1	3.9
	2	5.79	5.74	5.84	20.3	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-45MSD

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100

SAS No.: Q2100 SDG NO.: Q2100

Lab Sample ID: Q2100-04MSD

Date(s) Analyzed: 05/22/2025 05/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.79	6.74	6.84	22.0	3.7
	2	6.08	6.03	6.13	21.2	
4,4'-DDD	1	6.71	6.66	6.76	24.8	12.4
	2	5.93	5.88	5.98	21.9	
4,4'-DDT	1	7.02	6.97	7.07	18.4	1.6
	2	6.19	6.14	6.24	18.1	
Endrin aldehyde	1	6.92	6.87	6.97	21.5	3.8
	2	6.26	6.21	6.31	20.7	
Endosulfan sulfate	1	7.15	7.10	7.20	21.9	5.6
	2	6.48	6.43	6.53	20.7	
Methoxychlor	1	7.50	7.45	7.55	17.9	3.4
	2	6.76	6.71	6.81	17.3	
Endrin ketone	1	7.63	7.58	7.68	21.9	4.5
	2	6.99	6.94	7.04	22.9	
alpha-BHC	1	4.00	3.95	4.05	23.0	2.2
	2	3.39	3.34	3.44	22.5	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	23.0	3.1
	2	3.73	3.68	3.78	22.3	
Heptachlor	1	4.93	4.88	4.98	22.2	4.1
	2	4.08	4.03	4.13	21.3	
Aldrin	1	5.27	5.22	5.32	23.0	4
	2	4.37	4.32	4.42	22.1	
beta-BHC	1	4.52	4.47	4.57	22.2	2.2
	2	4.03	3.98	4.08	22.7	
delta-BHC	1	4.77	4.72	4.82	24.7	10.2
	2	4.26	4.21	4.31	22.3	
Heptachlor epoxide	1	5.69	5.64	5.74	22.6	3.6
	2	4.87	4.82	4.92	21.8	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-45MSD

Contract: CAMP02

Lab Code: CHEM **Case No.:** Q2100

SAS No.: Q2100

SDG NO.: Q2100

Lab Sample ID: Q2100-04MSD

Date(s) Analyzed: 05/22/2025 05/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1 **ID:** 0.32 (mm)

GC Column:(2): ZB-MR2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.08	6.03	6.13	22.5	4.5
	2	5.25	5.20	5.30	21.5	
gamma-Chlordane	1	5.95	5.90	6.00	22.5	3.2
	2	5.13	5.08	5.18	21.8	
alpha-Chlordane	1	6.03	5.98	6.08	22.5	3.6
	2	5.19	5.14	5.24	21.7	
4,4'-DDE	1	6.20	6.15	6.25	22.6	6.4
	2	5.38	5.33	5.43	21.2	
Dieldrin	1	6.35	6.30	6.40	22.5	3.6
	2	5.51	5.46	5.56	21.7	
Endrin	1	6.58	6.53	6.63	21.9	4.7
	2	5.79	5.74	5.84	20.9	