

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PD088720.D Time Analyzed: 11:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.707	6.608	6.808	56.410	50.000	12.8
4,4'-DDE	6.198	6.098	6.298	50.920	50.000	1.8
4,4'-DDT	7.023	6.923	7.123	42.850	50.000	-14.3
Aldrin	5.273	5.173	5.373	53.560	50.000	7.1
alpha-BHC	4.001	3.900	4.100	55.320	50.000	10.6
alpha-Chlordane	6.029	5.930	6.130	51.970	50.000	3.9
beta-BHC	4.517	4.417	4.617	52.540	50.000	5.1
Decachlorobiphenyl	9.077	8.977	9.177	49.910	50.000	-0.2
delta-BHC	4.766	4.666	4.866	57.170	50.000	14.3
Dieldrin	6.349	6.250	6.450	51.790	50.000	3.6
Endosulfan I	6.076	5.977	6.177	51.680	50.000	3.4
Endosulfan II	6.789	6.688	6.888	50.870	50.000	1.7
Endosulfan sulfate	7.152	7.052	7.252	49.160	50.000	-1.7
Endrin	6.575	6.477	6.677	49.640	50.000	-0.7
Endrin aldehyde	6.916	6.817	7.017	46.320	50.000	-7.4
Endrin ketone	7.632	7.533	7.733	49.560	50.000	-0.9
gamma-BHC (Lindane)	4.332	4.232	4.432	54.390	50.000	8.8
gamma-Chlordane	5.948	5.848	6.048	52.300	50.000	4.6
Heptachlor	4.931	4.832	5.032	52.660	50.000	5.3
Heptachlor epoxide	5.693	5.593	5.793	52.220	50.000	4.4
Methoxychlor	7.495	7.395	7.595	42.510	50.000	-15.0
Tetrachloro-m-xylene	3.551	3.451	3.651	54.790	50.000	9.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PD088720.D Time Analyzed: 11:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.931	5.833	6.033	50.560	50.000	1.1
4,4'-DDE	5.375	5.278	5.478	48.410	50.000	-3.2
4,4'-DDT	6.185	6.087	6.287	41.670	50.000	-16.7
Aldrin	4.369	4.271	4.471	51.080	50.000	2.2
alpha-BHC	3.394	3.294	3.494	52.800	50.000	5.6
alpha-Chlordane	5.192	5.094	5.294	49.090	50.000	-1.8
beta-BHC	4.026	3.927	4.127	50.280	50.000	0.6
Decachlorobiphenyl	8.074	7.977	8.177	45.050	50.000	-9.9
delta-BHC	4.263	4.164	4.364	51.650	50.000	3.3
Dieldrin	5.514	5.416	5.616	49.050	50.000	-1.9
Endosulfan I	5.248	5.150	5.350	44.660	50.000	-10.7
Endosulfan II	6.082	5.984	6.184	47.290	50.000	-5.4
Endosulfan sulfate	6.484	6.386	6.586	46.140	50.000	-7.7
Endrin	5.790	5.692	5.892	48.040	50.000	-3.9
Endrin aldehyde	6.260	6.162	6.362	44.580	50.000	-10.8
Endrin ketone	6.993	6.895	7.095	45.570	50.000	-8.9
gamma-BHC (Lindane)	3.731	3.631	3.831	51.950	50.000	3.9
gamma-Chlordane	5.127	5.029	5.229	49.630	50.000	-0.7
Heptachlor	4.084	3.985	4.185	49.830	50.000	-0.3
Heptachlor epoxide	4.874	4.775	4.975	49.900	50.000	-0.2
Methoxychlor	6.756	6.658	6.858	40.980	50.000	-18.0
Tetrachloro-m-xylene	2.882	2.782	2.982	53.980	50.000	8.0