

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PD088733.D Time Analyzed: 17:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.707	6.608	6.808	56.660	50.000	13.3
4,4'-DDE	6.198	6.098	6.298	50.690	50.000	1.4
4,4'-DDT	7.023	6.923	7.123	42.780	50.000	-14.4
Aldrin	5.273	5.173	5.373	57.200	50.000	14.4
alpha-BHC	3.999	3.900	4.100	59.800	50.000	19.6
alpha-Chlordane	6.029	5.930	6.130	53.120	50.000	6.2
beta-BHC	4.516	4.417	4.617	56.100	50.000	12.2
Decachlorobiphenyl	9.078	8.977	9.177	51.560	50.000	3.1
delta-BHC	4.765	4.666	4.866	61.300	50.000	22.6
Dieldrin	6.349	6.250	6.450	52.170	50.000	4.3
Endosulfan I	6.076	5.977	6.177	52.930	50.000	5.9
Endosulfan II	6.788	6.688	6.888	50.260	50.000	0.5
Endosulfan sulfate	7.152	7.052	7.252	47.820	50.000	-4.4
Endrin	6.575	6.477	6.677	51.430	50.000	2.9
Endrin aldehyde	6.917	6.817	7.017	48.240	50.000	-3.5
Endrin ketone	7.632	7.533	7.733	49.160	50.000	-1.7
gamma-BHC (Lindane)	4.331	4.232	4.432	58.430	50.000	16.9
gamma-Chlordane	5.947	5.848	6.048	53.490	50.000	7.0
Heptachlor	4.931	4.832	5.032	56.620	50.000	13.2
Heptachlor epoxide	5.692	5.593	5.793	54.000	50.000	8.0
Methoxychlor	7.496	7.395	7.595	42.370	50.000	-15.3
Tetrachloro-m-xylene	3.551	3.451	3.651	58.280	50.000	16.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PD088733.D Time Analyzed: 17:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.931	5.833	6.033	51.260	50.000	2.5
4,4'-DDE	5.375	5.278	5.478	50.960	50.000	1.9
4,4'-DDT	6.186	6.087	6.287	41.630	50.000	-16.7
Aldrin	4.368	4.271	4.471	54.460	50.000	8.9
alpha-BHC	3.394	3.294	3.494	56.540	50.000	13.1
alpha-Chlordane	5.191	5.094	5.294	51.670	50.000	3.3
beta-BHC	4.026	3.927	4.127	52.610	50.000	5.2
Decachlorobiphenyl	8.074	7.977	8.177	40.400	50.000	-19.2
delta-BHC	4.263	4.164	4.364	54.610	50.000	9.2
Dieldrin	5.514	5.416	5.616	50.340	50.000	0.7
Endosulfan I	5.247	5.150	5.350	44.260	50.000	-11.5
Endosulfan II	6.081	5.984	6.184	48.550	50.000	-2.9
Endosulfan sulfate	6.484	6.386	6.586	46.280	50.000	-7.4
Endrin	5.791	5.692	5.892	51.630	50.000	3.3
Endrin aldehyde	6.260	6.162	6.362	45.820	50.000	-8.4
Endrin ketone	6.992	6.895	7.095	45.000	50.000	-10.0
gamma-BHC (Lindane)	3.730	3.631	3.831	55.800	50.000	11.6
gamma-Chlordane	5.127	5.029	5.229	52.540	50.000	5.1
Heptachlor	4.084	3.985	4.185	53.360	50.000	6.7
Heptachlor epoxide	4.873	4.775	4.975	52.690	50.000	5.4
Methoxychlor	6.756	6.658	6.858	40.340	50.000	-19.3
Tetrachloro-m-xylene	2.881	2.782	2.982	56.290	50.000	12.6