

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG No.: Q2100

Instrument ID: BNA\_F Calibration Date/Time: 05/23/2025 11:12

Lab File ID: BF142527.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

| COMPOUND                    | RRF   | RRF040 | MIN<br>RRF | %D   | MAX%D |
|-----------------------------|-------|--------|------------|------|-------|
| 2-Fluorophenol              | 1.187 | 1.152  |            | -2.9 |       |
| Benzaldehyde                | 0.859 | 0.837  |            | -2.6 |       |
| Phenol-d6                   | 1.429 | 1.392  |            | -2.6 |       |
| Phenol                      | 1.608 | 1.583  |            | -1.6 | 20.0  |
| bis(2-Chloroethyl)ether     | 1.157 | 1.112  |            | -3.9 |       |
| 2-Chlorophenol              | 1.293 | 1.248  |            | -3.5 |       |
| 2-Methylphenol              | 1.010 | 0.995  |            | -1.5 |       |
| 2,2-oxybis(1-Chloropropane) | 1.944 | 1.901  |            | -2.2 |       |
| Acetophenone                | 0.443 | 0.427  |            | -3.6 |       |
| 3+4-Methylphenols           | 1.293 | 1.264  |            | -2.2 |       |
| n-Nitroso-di-n-propylamine  | 0.886 | 0.848  | 0.050      | -4.3 |       |
| Nitrobenzene-d5             | 0.367 | 0.350  |            | -4.6 |       |
| Hexachloroethane            | 0.507 | 0.478  |            | -5.7 |       |
| Nitrobenzene                | 0.331 | 0.318  |            | -3.9 |       |
| Isophorone                  | 0.613 | 0.580  |            | -5.4 |       |
| 2-Nitrophenol               | 0.177 | 0.169  |            | -4.5 | 20.0  |
| 2,4-Dimethylphenol          | 0.312 | 0.304  |            | -2.6 |       |
| bis(2-Chloroethoxy)methane  | 0.383 | 0.365  |            | -4.7 |       |
| 2,4-Dichlorophenol          | 0.283 | 0.272  |            | -3.9 | 20.0  |
| Naphthalene                 | 0.985 | 0.937  |            | -4.9 |       |
| 4-Chloroaniline             | 0.399 | 0.390  |            | -2.3 |       |
| Hexachlorobutadiene         | 0.192 | 0.178  |            | -7.3 | 20.0  |
| Caprolactam                 | 0.080 | 0.084  |            | 5.0  |       |
| 4-Chloro-3-methylphenol     | 0.291 | 0.281  |            | -3.4 | 20.0  |
| 2-Methylnaphthalene         | 0.620 | 0.586  |            | -5.5 |       |
| Hexachlorocyclopentadiene   | 0.387 | 0.374  | 0.050      | -3.4 |       |
| 2,4,6-Trichlorophenol       | 0.387 | 0.380  |            | -1.8 | 20.0  |
| 2-Fluorobiphenyl            | 1.490 | 1.386  |            | -7.0 |       |
| 2,4,5-Trichlorophenol       | 0.408 | 0.389  |            | -4.7 |       |
| 1,1-Biphenyl                | 1.552 | 1.477  |            | -4.8 |       |
| 2-Chloronaphthalene         | 1.146 | 1.093  |            | -4.6 |       |
| 2-Nitroaniline              | 0.331 | 0.326  |            | -1.5 |       |
| Dimethylphthalate           | 1.320 | 1.270  |            | -3.8 |       |
| Acenaphthylene              | 1.936 | 1.847  |            | -4.6 |       |
| 2,6-Dinitrotoluene          | 0.285 | 0.275  |            | -3.5 |       |
| 3-Nitroaniline              | 0.311 | 0.314  |            | 1.0  |       |
| Acenaphthene                | 1.182 | 1.124  |            | -4.9 | 20.0  |
| 2,4-Dinitrophenol           | 0.147 | 0.149  | 0.050      | 1.4  |       |
| 4-Nitrophenol               | 0.230 | 0.239  | 0.050      | 3.9  |       |

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG No.: Q2100

Instrument ID: BNA\_F Calibration Date/Time: 05/23/2025 11:12

Lab File ID: BF142527.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

| COMPOUND                   | RRF   | RRF040 | MIN<br>RRF | %D    | MAX%D |
|----------------------------|-------|--------|------------|-------|-------|
| Dibenzofuran               | 1.701 | 1.625  |            | -4.5  |       |
| 2,4-Dinitrotoluene         | 0.372 | 0.376  |            | 1.1   |       |
| Diethylphthalate           | 1.289 | 1.248  |            | -3.2  |       |
| 4-Chlorophenyl-phenylether | 0.646 | 0.620  |            | -4.0  |       |
| Fluorene                   | 1.314 | 1.263  |            | -3.9  |       |
| 4-Nitroaniline             | 0.282 | 0.300  |            | 6.4   |       |
| 4,6-Dinitro-2-methylphenol | 0.112 | 0.115  |            | 2.7   |       |
| n-Nitrosodiphenylamine     | 0.684 | 0.633  |            | -7.5  | 20.0  |
| 2,4,6-Tribromophenol       | 0.222 | 0.216  |            | -2.7  |       |
| 4-Bromophenyl-phenylether  | 0.236 | 0.214  |            | -9.3  |       |
| Hexachlorobenzene          | 0.262 | 0.242  |            | -7.6  |       |
| Atrazine                   | 0.182 | 0.183  |            | 0.5   |       |
| Pentachlorophenol          | 0.148 | 0.150  |            | 1.4   | 20.0  |
| Phenanthrene               | 1.069 | 1.006  |            | -5.9  |       |
| Anthracene                 | 1.091 | 1.036  |            | -5.0  |       |
| Carbazole                  | 0.933 | 0.921  |            | -1.3  |       |
| Di-n-butylphthalate        | 1.021 | 1.043  |            | 2.2   |       |
| Fluoranthene               | 0.999 | 0.980  |            | -1.9  | 20.0  |
| Pyrene                     | 1.866 | 1.802  |            | -3.4  |       |
| Terphenyl-d14              | 1.464 | 1.397  |            | -4.6  |       |
| Butylbenzylphthalate       | 0.516 | 0.553  |            | 7.2   |       |
| 3,3-Dichlorobenzidine      | 0.405 | 0.413  |            | 2.0   |       |
| Benzo (a) anthracene       | 1.336 | 1.324  |            | -0.9  |       |
| Chrysene                   | 1.200 | 1.134  |            | -5.5  |       |
| Bis(2-ethylhexyl)phthalate | 0.660 | 0.677  |            | 2.6   |       |
| Di-n-octyl phthalate       | 1.290 | 1.118  |            | -13.3 | 20.0  |
| Benzo (b) fluoranthene     | 1.184 | 1.171  |            | -1.1  |       |
| Benzo (k) fluoranthene     | 1.109 | 0.983  |            | -11.4 |       |
| Benzo (a) pyrene           | 1.116 | 1.065  |            | -4.6  | 20.0  |
| Indeno (1,2,3-cd) pyrene   | 1.501 | 1.373  |            | -8.5  |       |
| Dibenzo (a,h) anthracene   | 1.218 | 1.112  |            | -8.7  |       |
| Benzo (g,h,i) perylene     | 1.218 | 1.127  |            | -7.5  |       |
| 1,2,4,5-Tetrachlorobenzene | 0.574 | 0.550  |            | -4.2  |       |
| 1,4-Dioxane                | 0.475 | 0.471  |            | -0.8  | 20.0  |
| 2,3,4,6-Tetrachlorophenol  | 0.341 | 0.338  |            | -0.9  |       |

All other compounds must meet a minimum RRF of 0.010.