

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG No.: Q2100

Instrument ID: MSVOA_Y Calibration Date/Time: 05/22/2025 08:35

Lab File ID: VY022380.D Init. Calib. Date(s): 05/15/2025 05/15/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 11:47

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.522	0.512		-1.92	20
Chloromethane	1.097	1.093	0.1	-0.46	20
Vinyl Chloride	1.226	1.451		18.35	20
Bromomethane	0.998	1.202		20.44	20
Chloroethane	0.763	0.992		30.01	20
Trichlorofluoromethane	1.241	1.338		7.82	20
1,1,2-Trichlorotrifluoroethane	0.539	0.600		11.32	20
1,1-Dichloroethene	0.535	0.576		7.66	20
Acetone	0.102	0.096		-5.88	20
Carbon Disulfide	1.744	1.805		3.5	20
Methyl tert-butyl Ether	1.475	1.522		3.19	20
Methyl Acetate	0.323	0.304		-5.88	20
Methylene Chloride	0.662	0.614		-7.25	20
trans-1,2-Dichloroethene	0.590	0.630		6.78	20
1,1-Dichloroethane	1.106	1.155	0.1	4.43	20
Cyclohexane	1.105	1.072		-2.99	20
2-Butanone	0.171	0.154		-9.94	20
Carbon Tetrachloride	0.509	0.548		7.66	20
cis-1,2-Dichloroethene	0.681	0.741		8.81	20
Bromochloromethane	0.472	0.490		3.81	20
Chloroform	1.061	1.157		9.05	20
1,1,1-Trichloroethane	0.974	1.054		8.21	20
Methylcyclohexane	0.660	0.694		5.15	20
Benzene	1.457	1.548		6.25	20
1,2-Dichloroethane	0.394	0.396		0.51	20
Trichloroethene	0.362	0.385		6.35	20
1,2-Dichloropropane	0.351	0.362		3.13	20
Bromodichloromethane	0.492	0.514		4.47	20
4-Methyl-2-Pentanone	0.253	0.230		-9.09	20
Toluene	0.923	0.978		5.96	20
t-1,3-Dichloropropene	0.480	0.478		-0.42	20
cis-1,3-Dichloropropene	0.550	0.565		2.73	20
1,1,2-Trichloroethane	0.243	0.246		1.24	20
2-Hexanone	0.170	0.151		-11.18	20
Dibromochloromethane	0.315	0.330		4.76	20
1,2-Dibromoethane	0.226	0.236		4.43	20
Tetrachloroethene	0.462	0.502		8.66	20
Chlorobenzene	1.128	1.194	0.3	5.85	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG No.: Q2100

Instrument ID: MSVOA_Y Calibration Date/Time: 05/22/2025 08:35

Lab File ID: VY022380.D Init. Calib. Date(s): 05/15/2025 05/15/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 11:47

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	2.113	2.234		5.73	20
m/p-Xylenes	0.796	0.849		6.66	20
o-Xylene	0.753	0.791		5.05	20
Styrene	1.270	1.325		4.33	20
Bromoform	0.209	0.212	0.1	1.43	20
Isopropylbenzene	4.142	4.432		7	20
1,1,2,2-Tetrachloroethane	0.637	0.657	0.3	3.14	20
1,3-Dichlorobenzene	1.806	1.920		6.31	20
1,4-Dichlorobenzene	1.737	1.831		5.41	20
1,2-Dichlorobenzene	1.517	1.600		5.47	20
1,2-Dibromo-3-Chloropropane	0.103	0.104		0.97	20
1,2,4-Trichlorobenzene	0.876	0.937		6.96	20
1,2,3-Trichlorobenzene	0.743	0.779		4.84	20
1,2-Dichloroethane-d4	0.546	0.561		2.75	20
Dibromofluoromethane	0.298	0.318		6.71	20
Toluene-d8	1.222	1.298		6.22	20
4-Bromofluorobenzene	0.387	0.387		0	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.