

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG No.: Q2100

Instrument ID: MSVOA_Y Calibration Date/Time: 05/23/2025 14:34

Lab File ID: VY022407.D Init. Calib. Date(s): 05/15/2025 05/15/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 11:47

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.522	0.529		1.34	20
Chloromethane	1.097	1.312	0.1	19.6	20
Vinyl Chloride	1.226	1.569		27.98	20
Bromomethane	0.998	1.869		87.28	20
Chloroethane	0.763	1.174		53.87	20
Trichlorofluoromethane	1.241	1.446		16.52	20
1,1,2-Trichlorotrifluoroethane	0.539	0.598		10.95	20
1,1-Dichloroethene	0.535	0.580		8.41	20
Acetone	0.102	0.112		9.8	20
Carbon Disulfide	1.744	1.866		6.99	20
Methyl tert-butyl Ether	1.475	1.534		4	20
Methyl Acetate	0.323	0.277		-14.24	20
Methylene Chloride	0.662	0.610		-7.86	20
trans-1,2-Dichloroethene	0.590	0.629		6.61	20
1,1-Dichloroethane	1.106	1.138	0.1	2.89	20
Cyclohexane	1.105	1.069		-3.26	20
2-Butanone	0.171	0.165		-3.51	20
Carbon Tetrachloride	0.509	0.543		6.68	20
cis-1,2-Dichloroethene	0.681	0.745		9.4	20
Bromochloromethane	0.472	0.452		-4.24	20
Chloroform	1.061	1.137		7.16	20
1,1,1-Trichloroethane	0.974	1.014		4.11	20
Methylcyclohexane	0.660	0.714		8.18	20
Benzene	1.457	1.555		6.73	20
1,2-Dichloroethane	0.394	0.407		3.3	20
Trichloroethene	0.362	0.386		6.63	20
1,2-Dichloropropane	0.351	0.363		3.42	20
Bromodichloromethane	0.492	0.522		6.1	20
4-Methyl-2-Pentanone	0.253	0.241		-4.74	20
Toluene	0.923	0.988		7.04	20
t-1,3-Dichloropropene	0.480	0.487		1.46	20
cis-1,3-Dichloropropene	0.550	0.574		4.36	20
1,1,2-Trichloroethane	0.243	0.261		7.41	20
2-Hexanone	0.170	0.166		-2.35	20
Dibromochloromethane	0.315	0.339		7.62	20
1,2-Dibromoethane	0.226	0.243		7.52	20
Tetrachloroethene	0.462	0.463		0.22	20
Chlorobenzene	1.128	1.192	0.3	5.67	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG No.: Q2100

Instrument ID: MSVOA_Y Calibration Date/Time: 05/23/2025 14:34

Lab File ID: VY022407.D Init. Calib. Date(s): 05/15/2025 05/15/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 11:47

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	2.113	2.220		5.06	20
m/p-Xylenes	0.796	0.846		6.28	20
o-Xylene	0.753	0.797		5.84	20
Styrene	1.270	1.334		5.04	20
Bromoform	0.209	0.212	0.1	1.43	20
Isopropylbenzene	4.142	4.256		2.75	20
1,1,2,2-Tetrachloroethane	0.637	0.668	0.3	4.87	20
1,3-Dichlorobenzene	1.806	1.855		2.71	20
1,4-Dichlorobenzene	1.737	1.766		1.67	20
1,2-Dichlorobenzene	1.517	1.547		1.98	20
1,2-Dibromo-3-Chloropropane	0.103	0.101		-1.94	20
1,2,4-Trichlorobenzene	0.876	0.871		-0.57	20
1,2,3-Trichlorobenzene	0.743	0.748		0.67	20
1,2-Dichloroethane-d4	0.546	0.475		-13	20
Dibromofluoromethane	0.298	0.280		-6.04	20
Toluene-d8	1.222	1.060		-13.26	20
4-Bromofluorobenzene	0.387	0.334		-13.69	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.