

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Mahesh Dadoda	Review On	5/16/2025 11:49:17 AM				
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM				
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m		
STD. NAME	STD REF.#						
Tune/Reschk	VP133916						
Initial Calibration Stds	VP133918,VP133919,VP133920,VP133921,VP133922,VP133923						
CCC	VP133925						
Internal Standard/PEM	VP131783						
ICV/I.BLK	VP133924						
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022252.D	15 May 2025 07:52		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022253.D	15 May 2025 09:46		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022254.D	15 May 2025 10:16		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022255.D	15 May 2025 10:39		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022256.D	15 May 2025 11:02		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022257.D	15 May 2025 11:24		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022258.D	15 May 2025 11:47		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022259.D	15 May 2025 12:34		SY/MD	Ok
9	VSTDICV050	ICVVY051525	VY022260.D	15 May 2025 13:20		SY/MD	Ok,M
10	VY0515SBL01	VY0515SBL01	VY022261.D	15 May 2025 13:44		SY/MD	Ok
11	VY0515SBS01	VY0515SBS01	VY022262.D	15 May 2025 14:07		SY/MD	Ok,M
12	VY0515SBSD01	VY0515SBSD01	VY022263.D	15 May 2025 14:29		SY/MD	Ok,M
13	Q1984-01	OU4-PCS-TC-33-05072	VY022264.D	15 May 2025 15:12	vial-A	SY/MD	Ok
14	Q1984-03	OU4-PCS-TC-34-05072	VY022265.D	15 May 2025 15:35	vial-A	SY/MD	Ok
15	Q1984-05	OU4-PCS-TC-35-05072	VY022266.D	15 May 2025 15:59	vial-A	SY/MD	Ok
16	Q1984-07	OU4-TS-24-050725	VY022267.D	15 May 2025 16:22	vial-A Surrogate fail	SY/MD	ReRun
17	Q1984-09	OU4-TS-25-050725	VY022268.D	15 May 2025 16:46	vial-A Surrogate fail	SY/MD	Not Ok
18	Q1984-11	OU4-TS-26-050725	VY022269.D	15 May 2025 17:09	vial-A Surrogate fail	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Maresh Dadoda	Review On	5/16/2025 11:49:17 AM
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP133916		
Initial Calibration Stds	VP133918,VP133919,VP133920,VP133921,VP133922,VP133923		
CCC	VP133925		
Internal Standard/PEM	VP131783		
ICV/I.BLK	VP133924		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1984-13	OU4-TS-27-050725	VY022270.D	15 May 2025 17:33	vial-A Surrogate fail	SY/MD	Not Ok
20	Q1984-15	OU4-TS-28-050725	VY022271.D	15 May 2025 17:56	vial-A Surrogate fail	SY/MD	ReRun
21	Q1982-04	TP-4	VY022272.D	15 May 2025 18:19	vial-A	SY/MD	Ok
22	Q1982-05	TP-5	VY022273.D	15 May 2025 18:43	vial-A	SY/MD	Ok
23	Q1982-06	TP-6	VY022274.D	15 May 2025 19:06	vial-A	SY/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VY022275.D	15 May 2025 19:52		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052225

Review By	John Carlone	Review On	5/23/2025 11:05:57 AM
Supervise By	Maresh Dadoda	Supervise On	5/23/2025 4:07:06 PM
SubDirectory	VY052225	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134001
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134002,VP134003 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022379.D	22 May 2025 08:05		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022380.D	22 May 2025 08:35		SY/MD	Ok,M
3	VY0522SBL01	VY0522SBL01	VY022381.D	22 May 2025 09:07		SY/MD	Ok
4	VY0522SBS01	VY0522SBS01	VY022382.D	22 May 2025 09:36		SY/MD	Ok,M
5	VY0522SBSD01	VY0522SBSD01	VY022383.D	22 May 2025 09:59		SY/MD	Ok,M
6	Q2087-03RE	NB-08-05202025RE	VY022384.D	22 May 2025 10:38	vial-B ,Internal Standard Fail	SY/MD	Confirms
7	Q2101-03	TP-1-MHE-VOC	VY022385.D	22 May 2025 11:01	vial-B	SY/MD	Ok
8	Q2097-11RE	RT3419RE	VY022386.D	22 May 2025 11:25	vial-B ,Internal Standard Fail	SY/MD	Confirms
9	Q2102-01	LAW-25-0077	VY022387.D	22 May 2025 11:48	vial-B	SY/MD	Ok
10	Q2104-01RE	TR-06-052125RE	VY022388.D	22 May 2025 12:11	vial-B ,Internal Standard Fail; Surrogate Fail	SY/MD	Confirms
11	Q2074-01	TP-12	VY022389.D	22 May 2025 12:35	vial-A	SY/MD	Ok
12	Q2074-02	TP-7	VY022390.D	22 May 2025 12:58	vial-A	SY/MD	Ok
13	Q2074-03	TP-15	VY022391.D	22 May 2025 13:22	vial-A	SY/MD	Ok
14	Q2074-04	TP-20	VY022392.D	22 May 2025 13:45	vial-A	SY/MD	Ok
15	Q2074-05	TP-38	VY022393.D	22 May 2025 14:09	vial-A	SY/MD	Ok
16	Q2074-06	TP-19	VY022394.D	22 May 2025 14:32	vial-A	SY/MD	Ok
17	Q2074-07	TP-40	VY022395.D	22 May 2025 14:55	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052225

Review By	John Carlone	Review On	5/23/2025 11:05:57 AM
Supervise By	Maresh Dadoda	Supervise On	5/23/2025 4:07:06 PM
SubDirectory	VY052225	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134001		
CCC	VP134002,VP134003		
Internal Standard/PEM	VP133934		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2074-08	TP-18	VY022396.D	22 May 2025 15:19	vial-A	SY/MD	Ok
19	Q2100-01	TP-16	VY022397.D	22 May 2025 15:42	vial-A ,Internal Standard Fail	SY/MD	ReRun
20	Q2100-02	TP-22	VY022398.D	22 May 2025 16:06	vial-A	SY/MD	Ok
21	Q2100-03	TP-21	VY022399.D	22 May 2025 16:29	vial-A ,Not purge	SY/MD	Not Ok
22	Q2100-04	TP-45	VY022400.D	22 May 2025 16:53	vial-A	SY/MD	Ok
23	Q2112-01	EO-03-05222025	VY022401.D	22 May 2025 17:16	vial-A ,Internal Standard Fail	SY/MD	ReRun
24	Q2109-03	TP-02-MHF-VOC	VY022402.D	22 May 2025 17:40	vial-A	SY/MD	Ok
25	Q2075-06	SS-MW1-11.5	VY022403.D	22 May 2025 18:03	vial-A ,Need MeOH	SY/MD	Dilution
26	VIBLK	VIBLK	VY022404.D	22 May 2025 18:26		SY/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VY022405.D	22 May 2025 18:49		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052325

Review By	Mahesh Dadoda	Review On	5/27/2025 4:51:38 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/27/2025 4:53:09 AM		
SubDirectory	VY052325	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y051525S.M
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134014			
CCC		VP134015,VP135016			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022406.D	23 May 2025 11:40		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022407.D	23 May 2025 14:34		SY/MD	Ok,M
3	VY0523SBL01	VY0523SBL01	VY022408.D	23 May 2025 15:26		SY/MD	Ok
4	VY0523SBS01	VY0523SBS01	VY022409.D	23 May 2025 16:28		SY/MD	Ok,M
5	VY0523SBSD01	VY0523SBSD01	VY022410.D	23 May 2025 16:50		SY/MD	Ok,M
6	Q2118-11	BP-TPB-182-GW-500-5	VY022411.D	23 May 2025 17:14	Surrogate Fail	SY/MD	ReRun
7	Q2112-01RE	EO-03-05222025RE	VY022412.D	23 May 2025 17:37	Internal Standard Fail	SY/MD	Confirms
8	Q2100-03	TP-21	VY022413.D	23 May 2025 18:01	vial-B	SY/MD	Ok
9	Q2127-11	COMP-2	VY022414.D	23 May 2025 18:24	Internal Standard Fail	SY/MD	ReRun
10	Q2127-01	COMP-1	VY022415.D	23 May 2025 18:47		SY/MD	Ok
11	Q2100-01	TP-16	VY022416.D	23 May 2025 19:11	vial-B	SY/MD	Ok
12	VIBLK	VIBLK	VY022417.D	23 May 2025 19:34		SY/MD	Ok
13	VY0523SBS02	VY0523SBS02	VY022418.D	23 May 2025 19:57	Not required	SY/MD	Not Ok
14	VSTDCCC050	VSTDCCC050EC	VY022419.D	23 May 2025 21:05		SY/MD	Ok,M

M : Manual Integration