

Surrogate Summary

SDG No.: **Q2100**

Client: **CDM Smith**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2100-01	TP-16	1,2-Dichloroethane-d4	50	55.5	111	63	155
		Dibromofluoromethane	50	51.9	104	70	134
		Toluene-d8	50	48.8	98	74	123
		4-Bromofluorobenzene	50	40.9	82	38	136
Q2100-02	TP-22	1,2-Dichloroethane-d4	50	46.6	93	63	155
		Dibromofluoromethane	50	48.6	97	70	134
		Toluene-d8	50	48.7	97	74	123
		4-Bromofluorobenzene	50	38.3	77	38	136
Q2100-03	TP-21	1,2-Dichloroethane-d4	50	48.8	98	63	155
		Dibromofluoromethane	50	50.0	100	70	134
		Toluene-d8	50	47.9	96	74	123
		4-Bromofluorobenzene	50	37.3	75	38	136
Q2100-04	TP-45	1,2-Dichloroethane-d4	50	50.5	101	63	155
		Dibromofluoromethane	50	51.5	103	70	134
		Toluene-d8	50	48.7	97	74	123
		4-Bromofluorobenzene	50	39.7	79	38	136
VY0522SBL01	VY0522SBL01	1,2-Dichloroethane-d4	50	48.9	98	63	155
		Dibromofluoromethane	50	49.8	100	70	134
		Toluene-d8	50	48.7	97	74	123
		4-Bromofluorobenzene	50	39.5	79	38	136
VY0522SBS01	VY0522SBS01	1,2-Dichloroethane-d4	50	48.4	97	63	155
		Dibromofluoromethane	50	50.3	101	70	134
		Toluene-d8	50	50.3	101	74	123
		4-Bromofluorobenzene	50	47.7	95	38	136
VY0522SBSD01	VY0522SBSD01	1,2-Dichloroethane-d4	50	50.4	101	63	155
		Dibromofluoromethane	50	51.0	102	70	134
		Toluene-d8	50	49.9	100	74	123
		4-Bromofluorobenzene	50	47.1	94	38	136
VY0523SBL01	VY0523SBL01	1,2-Dichloroethane-d4	50	56.1	112	63	155
		Dibromofluoromethane	50	56.7	113	70	134
		Toluene-d8	50	54.7	109	74	123
		4-Bromofluorobenzene	50	44.0	88	38	136
VY0523SBS01	VY0523SBS01	1,2-Dichloroethane-d4	50	50.7	101	63	155
		Dibromofluoromethane	50	53.6	107	70	134
		Toluene-d8	50	53.5	107	74	123
		4-Bromofluorobenzene	50	48.8	98	38	136