

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2100 SAS No.: Q2100 SDG NO.: Q2100
 Lab File ID: VY022407.D Date Analyzed: 05/23/2025
 Instrument ID: MSVOA_Y Time Analyzed: 14:34
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	221240	7.72	368290	8.62	319713	11.42
UPPER LIMIT	442480	8.219	736580	9.122	639426	11.92
LOWER LIMIT	110620	7.219	184145	8.122	159857	10.92
EPA SAMPLE NO.						
TP-16	303749	7.71	546937	8.62	451641	11.42
TP-21	340717	7.71	608366	8.62	471350	11.42
VY0523SBL01	328707	7.71	590839	8.62	465211	11.42
VY0523SBS01	220484	7.71	365163	8.62	306453	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.