

Report of Analysis

Client:	PSEG	Date Collected:	05/19/25
Project:	PSEG East Edison Test Pits	Date Received:	05/20/25
Client Sample ID:	TP-1-MHE	SDG No.:	Q2101
Lab Sample ID:	Q2101-04	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024802.D	1	05/23/25 11:16	05/23/25 21:13	PB168156

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	0.050	U	0.013	0.050	mg/L
106-46-7	1,4-Dichlorobenzene	0.050	U	0.0053	0.050	mg/L
95-48-7	2-Methylphenol	0.050	U	0.011	0.050	mg/L
65794-96-9	3+4-Methylphenols	0.10	U	0.011	0.10	mg/L
67-72-1	Hexachloroethane	0.050	U	0.0065	0.050	mg/L
98-95-3	Nitrobenzene	0.050	U	0.0076	0.050	mg/L
87-68-3	Hexachlorobutadiene	0.050	U	0.0054	0.050	mg/L
88-06-2	2,4,6-Trichlorophenol	0.050	U	0.0051	0.050	mg/L
95-95-4	2,4,5-Trichlorophenol	0.050	U	0.0062	0.050	mg/L
121-14-2	2,4-Dinitrotoluene	0.050	U	0.012	0.050	mg/L
118-74-1	Hexachlorobenzene	0.050	U	0.0052	0.050	mg/L
87-86-5	Pentachlorophenol	0.10	U	0.016	0.10	mg/L
SURROGATES						
367-12-4	2-Fluorophenol	116		10 - 139	77%	SPK: 150
13127-88-3	Phenol-d6	98.2		10 - 134	65%	SPK: 150
4165-60-0	Nitrobenzene-d5	73.4		49 - 133	73%	SPK: 100
321-60-8	2-Fluorobiphenyl	73.6		52 - 132	74%	SPK: 100
118-79-6	2,4,6-Tribromophenol	161		44 - 137	108%	SPK: 150
1718-51-0	Terphenyl-d14	90.5		48 - 125	90%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	93400	7.652			
1146-65-2	Naphthalene-d8	358000	10.422			
15067-26-2	Acenaphthene-d10	239000	14.292			
1517-22-2	Phenanthrene-d10	522000	17.086			
1719-03-5	Chrysene-d12	580000	21.516			
1520-96-3	Perylene-d12	674000	24.804			

Report of Analysis

Client:	PSEG	Date Collected:	05/19/25
Project:	PSEG East Edison Test Pits	Date Received:	05/20/25
Client Sample ID:	TP-1-MHE	SDG No.:	Q2101
Lab Sample ID:	Q2101-04	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Level :	LOW
		Final Vol:	1000
		Test:	TCLP BNA
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024802.D	1	05/23/25 11:16	05/23/25 21:13	PB168156

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products