

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2105

Project: Monthly 2025

Run No.: LB135878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135878BS							
TPH	mg/L	20.0	16.8		84	1	78-114	05/22/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2105

Project: Monthly 2025

Run No.: LB135878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135878BSD							
TPH	mg/L	20.0	16.9		84	1	78-114	05/22/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2105

Project: Monthly 2025

Run No.: LB135887

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135887BS							
BOD5	mg/L	198	191		96	1	84.6-115.4	05/22/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2105

Project: Monthly 2025

Run No.: LB135909

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135909BS							
TSS	mg/L	550	532		97	1	90-110	05/23/2025