

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2106
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2102-03
Client ID:	LAW-25-0078DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Flash Point	o F	+/-2	>212.0		>212.0		1	0		05/23/2025
TSS	mg/L	+/-5	138		139		1	0.51		05/23/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: COMPDUP	SDG No.: Q2106 Sample ID: Q2105-02 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	1020		1030		1	1.13		05/22/2025