

### Laboratory Control Sample Summary

**Client:** ENTACT

**SDG No.:** Q2106

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**Run No.:** LB135887

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB135887BS |            |        |                 |            |                 |                     |               |
| BOD5      | mg/L       | 198        | 191    |                 | 96         | 1               | 84.6-115.4          | 05/22/2025    |

### Laboratory Control Sample Summary

**Client:** ENTACT

**SDG No.:** Q2106

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**Run No.:** LB135909

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB135909BS |            |        |                 |            |                 |                     |               |
| TSS       | mg/L       | 550        | 532    |                 | 97         | 1               | 90-110              | 05/23/2025    |