



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q2106
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2106 **SAS No.:** Q2106
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.24	0.2	118	70 - 130	CV	05/27/2025	11:56	LB135915
CRI01	Cadmium	5.78	6.0	96	65 - 135	P	05/28/2025	12:53	LB135935
	Copper	20.3	20.0	101	65 - 135	P	05/28/2025	12:53	LB135935
	Lead	12.8	12.0	107	65 - 135	P	05/28/2025	12:53	LB135935
	Nickel	38.0	40.0	95	65 - 135	P	05/28/2025	12:53	LB135935
	Zinc	43.8	40.0	109	65 - 135	P	05/28/2025	12:53	LB135935