



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q2106</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q2106</u>	SAS No.:	<u>Q2106</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB17	Mercury	0.20	+/-0.20	U	0.20	CV	05/27/2025	11:49	LB135915

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2106

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2106

SAS No.: Q2106

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	05/27/2025	11:53	LB135915
CCB66	Mercury	0.20	+/-0.20	U	0.20	CV	05/27/2025	12:21	LB135915
CCB67	Mercury	0.20	+/-0.20	U	0.20	CV	05/27/2025	12:49	LB135915
CCB68	Mercury	0.20	+/-0.20	U	0.20	CV	05/27/2025	13:02	LB135915

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2106

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2106

SAS No.: Q2106

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	12:45	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	12:45	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	12:45	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	12:45	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	12:45	LB135935

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q2106							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q2106	SAS No.: Q2106						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	13:26	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	13:26	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	13:26	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	13:26	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	13:26	LB135935
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	14:10	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	14:10	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	14:10	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	14:10	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	14:10	LB135935
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	14:56	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	14:56	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	14:56	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	14:56	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	14:56	LB135935
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	15:49	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	15:49	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	15:49	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	15:49	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	15:49	LB135935
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	16:43	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	16:43	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	16:43	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	16:43	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	16:43	LB135935
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	18:05	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	18:05	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	18:05	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	18:05	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	18:05	LB135935
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	18:55	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	18:55	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	18:55	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	18:55	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	18:55	LB135935
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	19:41	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	19:41	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	19:41	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	19:41	LB135935

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2106</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2106</u>	SAS No.: <u>Q2106</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	19:41	LB135935
CCB09	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	20:26	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	20:26	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	20:26	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	20:26	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	20:26	LB135935
CCB10	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	21:23	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	21:23	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	21:23	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	21:23	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	21:23	LB135935
CCB11	Cadmium	6.00	+/-6.00	U	6.00	P	05/28/2025	22:19	LB135935
	Copper	20.0	+/-20.0	U	20.0	P	05/28/2025	22:19	LB135935
	Lead	12.0	+/-12.0	U	12.0	P	05/28/2025	22:19	LB135935
	Nickel	40.0	+/-40.0	U	40.0	P	05/28/2025	22:19	LB135935
	Zinc	40.0	+/-40.0	U	40.0	P	05/28/2025	22:19	LB135935

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q2106</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q2106</u>	SAS No.:	<u>Q2106</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
