

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** Q2106
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2106 **SAS No.:** Q2106
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-0.014	1.0	1	-5	7	05/28/2025	13:03	LB135935
	Copper	-5.04	2.0	252	-18	22	05/28/2025	13:03	LB135935
	Lead	0.62			-12	12	05/28/2025	13:03	LB135935
	Nickel	1.34	2.0	67	-38	42	05/28/2025	13:03	LB135935
	Zinc	-0.72			-40	40	05/28/2025	13:03	LB135935
ICSAB01	Cadmium	1020	972	105	826	1120	05/28/2025	13:08	LB135935
	Copper	517	511	101	434	588	05/28/2025	13:08	LB135935
	Lead	50.7	49.0	104	37	61	05/28/2025	13:08	LB135935
	Nickel	1020	954	107	810	1100	05/28/2025	13:08	LB135935
	Zinc	1070	952	112	809	1095	05/28/2025	13:08	LB135935