

metals
- 5a -
MATRIX SPIKE SUMMARY

client: ENTACT **level:** low **sdg no.:** Q2106
contract: ENTA05 **lab code:** CHEM **case no.:** Q2106 **sas no.:** Q2106
matrix: Water **sample id:** Q2093-14 **client id:** WATER-COMPMS
Percent Solids for Sample: NA **Spiked ID:** Q2093-14MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	97.3		3.00	U	100	97		P
Copper	ug/L	75 - 125	153		3.37	J	150	100		P
Lead	ug/L	75 - 125	474		3.19	J	500	94		P
Nickel	ug/L	75 - 125	253		5.05	J	250	99		P
Zinc	ug/L	75 - 125	119		79.7		100	40	N	P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: ENTACT **level:** low **sdg no.:** Q2106
contract: ENTA05 **lab code:** CHEM **case no.:** Q2106 **sas no.:** Q2106
matrix: Water **sample id:** Q2093-14 **client id:** WATER-COMPMSD
Percent Solids for Sample: NA **Spiked ID:** Q2093-14MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	102		3.00	U	100	102		P
Copper	ug/L	75 - 125	162		3.37	J	150	105		P
Lead	ug/L	75 - 125	495		3.19	J	500	98		P
Nickel	ug/L	75 - 125	265		5.05	J	250	104		P
Zinc	ug/L	75 - 125	126		79.7		100	47	N	P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: ENTACT **level:** low **sdg no.:** Q2106
contract: ENTA05 **lab code:** CHEM **case no.:** Q2106 **sas no.:** Q2106
matrix: Water **sample id:** Q2102-03 **client id:** LAW-25-0078MS
Percent Solids for Sample: NA **Spiked ID:** Q2102-03MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	2.71		0.20	U	4.0	68	N	CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: ENTACT **level:** low **sdg no.:** Q2106
contract: ENTA05 **lab code:** CHEM **case no.:** Q2106 **sas no.:** Q2106
matrix: Water **sample id:** Q2102-03 **client id:** LAW-25-0078MSD
Percent Solids for Sample: NA **Spiked ID:** Q2102-03MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	2.64		0.20	U	4.0	66	N	CV